

Towards calibration transfer for quantitative analysis of nuclear materials by LIBS

E. Rollin, J. Picard, D. Cardona, O. Musset, J.-B. Sirven

► To cite this version:

E. Rollin, J. Picard, D. Cardona, O. Musset, J.-B. Sirven. Towards calibration transfer for quantitative analysis of nuclear materials by LIBS. 14th European Workshop on Laser Ablation 2018, Jun 2018, Pau, France. cea-02338859

HAL Id: cea-02338859 https://cea.hal.science/cea-02338859v1

Submitted on 21 Feb 2020 $\,$

HAL is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers. L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.

DE LA RECHERCHE À L'INDUSTRIE

ceaden



TOWARDS CALIBRATION TRANSFER FOR QUANTITATIVE ANALYSIS OF NUCLEAR MATERIALS BY LIBS

A unified approach for the characterization of the ablation plasma in various matrices

Édouard ROLLIN^A Jessica PICARD^B Danièle CARDONA^B Olivier MUSSET^C Jean-Baptiste SIRVEN^A

 ^A CEA Saclay, Department of Physical Chemistry (DPC), SEARS, LANIE
 ^B CEA Valduc
 ^C Université de Bourgogne Franche-Comté, ICB

Edouard.rollin@cea.fr

14TH EUROPEAN WORKSHOP ON LASER ABLATION

JUNE 27 PAU, FRANCE

Ceaden

A NEED FOR ONLINE ANALYSIS

Some elements of context :

CEA Valduc :

- <u>Chemical analysis</u>: quantification of metal impurities for quality control of nuclear metal materials from production lines...
- <u>Today's methods</u>: offline analysis;
 - Sometimes long delays between sampling and results;
 Time consuming.

 \rightarrow A first step of **online** quality pre-control may simplify the processes.







Perks of LIBS for this purpose

- Almost no sample preparation;
- Can be done remotely;
- Fast.
 - \rightarrow Optimization of production time;
 - \rightarrow Less irradiation risks;
 - → Reduction of wastes.

CALIBRATION TRANSFER

Impossibility to achieve an analytical development directly with nuclear materials:

- Lack of calibrated samples;
- Complexity of U and Pu spectroscopy;
- Security and wastes constraints.



Need for a general method developed from surrogate materials.

Three steps:

Ceaden

- Identification of surrogate materials;
- Analytical development on these surrogate materials;
- Calibration transfer to nuclear materials

Previous PhD Thesis: Jessica Picard (2012-2015)



But how can we achieve this calibration transfer ?



Ceaden

MATRIX EFFECTS

The matrix effects in a nutshell



→ Need to evaluate ablated mass, plasma temperature and electronic density in various matrices to correct the calibration curves.

EWLA, JUNE 27th, 2018 | PAGE 4

Number of analyte atoms in gas phase in the plasma: includes m₀ and N_a



Measure them in different matrices in order to:

→ Use them to normalize the calibrations and correct the matrix effects between various metal matrices.

→ Find relationships between them and the physical properties of the matrices in order to acquire a better understanding of the matrix effects.

Aim: quantify impurities in all kinds of metal matrices using corrected calibrations

Ceaden

EXPERIMENTAL SETUP



MobiLIBS system (IVEA Solution)

- Laser : Nd:YAG 266nm, 4ns and 9mJ pulses, 50µm spots
- Spectrometer : Horiba THR 1000 monochromator
- ICCD Andor iStar
- Certified metal standard samples





MEASUREMENT METHODS

Important hypothesis : Local Thermodynamic Equilibrium (LTE)

→ Legitimates the use of spectroscopic methods.

T_e : Boltzmann-plot

The intensity ratio between two lines of different energy levels is related to plasma temperature.

The hotter the plasma,

the more intense the lines of high energy.



<u>N_e : Stark broadening</u>

The width of a line is related, among other things, to plasma electronic density.

The denser the plasma,

the wider the emission lines



Ceaden A UNIFIED METH

A UNIFIED METHOD FOR TEMPERATURE MEASUREMENTS IN VARIOUS MATRICES



<u>ceaden</u>

A UNIFIED METHOD FOR TEMPERATURE MEASUREMENTS **IN VARIOUS MATRICES**



EWLA, JUNE 27th, 2018 | PAGE 9



Lots of spectral interferences with elements of the matrices especially Ti and Ni

→ Trade-off between the number of measured lines and the precision of their measurement in every matrix.

Either we keep the best lines for each matrix...



<u>Uncertainty</u> = ± one standard deviation of experimental repetability



Lots of spectral interferences with elements of the matrices especially Ti and Ni

→ Trade-off between the number of measured lines and the precision of their measurement in every matrix.

Or we keep only the lines that are ok for every matrix...



Ceaden MEASURING T_E IN VARIOUS MATRICES

The experimental procedure in a nutshell

Accumulation of laser shots... On several samples... Of several matrices... Aaaaaaaannnnddddd... Even more laser shots

. . .





RECHERCHE À L'INDUSTRI cea den

Not really significant for

Ni or Ti, though T_e in Ni

seems higher

RESULTS

Plasma temperature indicator in several matrices T (K) 7225 K 6904 K σ= 362 K 8000 σ= 600 K 6541 K 6895 K σ=121 K σ= 80 K 7500 7000 6500 6000 6655 K 6627 K 6489 K 6397 K 6806 K 6958 K 6922 K 7481 K 6970 K 6480 K 7329 K 5500 5000 Aluminum Nickel Copper Titanium **Small but significant**

variations between

Al and Cu

4 matrices 11 samples

Error bars = ± one standard deviation of experimental repetability

 σ = standard deviation for different samples of a same matrix

Temperature variations in a same matrix: different compositions, self-absorption...



Calibrations from previous works : normalized with ablated mass



RESULTS

Boltzmann Equation :



The higher the temperature, the higher the slope.

 \rightarrow In agreement with our temperature indicator

Need a whole new set of experiments with simultanous measurements of m_0 , T_e and N_e to make sure of the possible correction of matrix effects.

CONCLUSION AND FUTURE WORKS

What we found:

cea den

- We can measure an indicator for plasma temperature using integrated spectra. This indicator depends greatly on the matrix but only slightly on the precise composition of the sample.
- This indicator can help us correct the matrix effect.



What limits us:

No control over the composition of the samples.

→ Spectral interferences with other impurities, self-absorption...

What we will try:

- Find a way to work with pure metals or more controlled samples.
- Find an indicator for the electron density.



Thank you for your attention !

Any questions ?

Edouard.rollin@cea.fr



Commissariat à l'énergie atomique et aux énergies alternatives Centre de Saclay | 91191 Gif-sur-Yvette Cedex T. +33 (0)1 69 08 71 84

 T. +33 (0)1 69 08 71 84
 DPC

 Etablissement public à caractère industriel et commercial | RCS Paris B 775 685 019
 SEARS

DE LA RECHERCHE À L'INDUSTRIE

Ceaden



Appendix

DE LA RECHERCHE À L'INDUSTRIE



		Al						Cu						Ni			Steel				Ti			
	Samples	51XG00H2	637C/01	01002	311/01	124/04	35831	63698	CT1	CT2	СТЗ	СТ4 (CT5	B7011	B7004	F294	284-2	BSCA316-3	ECRM 289-1	BS85D	BST81	IA271A	BST22	BSTSU-1
Element	Molar mass				,																			
%massique	inorar mass																							
Ag	107 87					0.019			0.057	0 042	0.026	0.016	0.0062				1							
ΔΙ	26 982	98 3517	95 2308	81 4136	98 52	99.0654	97 4932	95.08	0.037	0.042	0.020	0.010	0.0002	12	16	4 0.0	0.002	7 0.004	0 199	0 13	0.0664	5 28	0.004	6.11
Δs	74 922	0.005	55.2500	01.4150	50.52	55.0054	57.4552	55.00	0.32	0.22	0.11	0.05	0.0056	1.2	. 1.0	- 0.0.	0.0027	0.004	0.0056	0.006	0.0004	5.20	0.004	0.11
B	10.81	0.005		0.0002		0.0002			0.024	0.033	0.0003	0.0042	0.0050				0.0026	0.0005	0.0030	1 0.0006	0.0082			0.01
Bo	9 0122			0.0002		0.0002			0.024	0.055	0.00055	0.0042	0.011				0.0020	0.0005	0.004	+ 0.0000	0.0002			0.01
DC Di	208 08	0.011				0.0015			0.019	0.012	0.0067	0.0042	0.0011											
C	12 011	0.011				0.0151			0.018	0.013	0.0007	0.0043	0.0011	0.095	0.0	1	0.0201	0.018	0.0490	0.049	0.0161	0.026	0.005	0.062
	12.011					0 0003								0.035	0.0	1	0.0201	0.018	0.048	0.043	0.0101	0.020	0.005	0.002
Cd	112 /1	0.0021				0.0005												0.005		0.0004				
Co	112.41	0.0031				0.021																		
Ce	E9 022					0.0451			0.051	0.022	0.012	0.011	0.0061	17	0.0	c 0.5	1 0.0525	0.22	0.065	0.07	0.0205			0.061
Cr	50.955	0.0174	0.2514	0.062	0.005	0.0201	0.0022	0.03	0.031	0.055	0.015	0.011	0.0001	10.7	10	6 0.0	1 16 911	17.40	14.63	0.97	0.0393	0.016	1 22	0.001
Ci	51.990	0.0174	0.5514	0.002	0.095	0.0509	0.0025	0.03	07.766	09.441	00 05027	00 7725	00 4101	19.7	19.	0.04	1 0.01	17.49	14.03	5 17.09	0.0294	0.010	1.22	0.07
Cu Fo	03.540	0.111	0.379	4.33	0.044	0.0533	0.005	0.4	97.700	98.441	98.95037	98.7735	98.4101	0.065	0.1	4 0.1	0.183	0.43	EE 2200	0.45	0.0244	0.004	0.04	0.073
Fe	55.845	0.350	0.377	0.076	0.21	0.145	0.558	0.81	0.17	0.1	0.083	0.045	0.016	0.34	0.9	/ 0.00	5 67.490	00.1381	55.3205	9 07.0372	0.1144	0.31	1.19	0.22
	1.000	0.011	0.0405			0.0265															0.0055	0.015	·'	
Ga	59.723	0.011	0.0495			0.0365																		
Ge	/2.03					0.0207																		
La	138.91	0.000	1 400	1.00	0.24	0.0307	0.200	1.20								0.07							0.000	0.005
IVIg	24.305	0.093	1.468	1.66	0.34	0.0558	0.208	1.36						0.10		0.0.			4.044		0.0404	0.000	0.003	0.005
Min	54.938	0.154	0.304	0.305	0.42	0.054	0.108	0.65	9					0.16	0.0	8 0.1	1 1.745	1.6	1.016	1.69	0.0404	0.002	2.02	0.065
IVIO	95.95													0.2	0.0	6	2.111	2.04	1.102	2 0.59	0.0279	0.011	1.15	0.067
N	14.007																0.0151	0.058		0.016	0.0037	0.012	0.01	0.02
ND	92.906	0.440	0.4020	0.000	0.05	0.0500	0.0575			0.00	0.43	0.040	0.000	50.47			10.75	0.031		0.065	0.0191	0.005	0.01	0.068
NI	58.693	0.119	0.1026	0.203	0.05	0.0592	0.0575		0.48	0.29	0.12	0.049	0.0095	58.17	/4.6	4 98.9	9 10.72	2 11.26	24.68	3 10.03	0.009	0.035	0.008	0.073
0	15.999																0.0099	0.0026		0.0014	0.0669	0.16	0.09	0.191
P	30.974								0.082	0.059	0.038	0.02	0.0059				0.0258	3 0.027	0.0114	4 0.025		0.002		0.009
Pb	207.2	0.114	0.02	0.0517	0.042	0.0232	0.08		0.013	0.086	0.31	0.88	1.48						0.0008	3 0.0004	0.0000			0.45
Pa	106.42																				0.0398			0.15
Ru	101.07								0.054	0.020	0.013	0.000	0.000				0.000		0.000		0.031	0.000	0.000	0.07
S	32.06	0.0005		0.0400		0.0050			0.054	0.036	0.012	0.006	0.0024				0.0237	0.024	0.002	0.024		0.002	0.002	0.008
SD	127.76	0.0265		0.0403		0.0252			0.33	0.24	0.11	0.049	0.01					0.0018	0.001:	3 0.001				
Se	/8.9/1	0.05	1.500	10.04	0.055	0.405			0.062	0.041	0.018	0.011	0.0065	1.02			0.50		0.53		0.0474	0.004	0.00	0.057
SI	28.085	0.35	1.569	10.94	0.055	0.105	1.1	1.33						1.02	0.1	/ 0.09	0.53	0.49	0.53	1 0.55	0.04/4	0.021	0.02	0.057
Sn -	118.71	0.0065	0.0022	0.0242		0.0208	0.051		0.24	0.14	0.07	0.025	0.007				0.0047	0.01	0.111	1	0.0155	2.49	0.019	0.076
Та	180.95																							0.079
Те	127.6								0.053	0.036	0.022	0.011	0.0064			_								
Ti	47.867	0.128	0.0648	0.2301	0.09	0.0505	0.166	0.06	5					2.05	2.6	3 0.03	3 0.191	L 0.033	2.01	1 0.48	99.3236	91.511	93.189	88.184
V	50.942	0.0148	0.0291	0.125	0.059	0.0449											0.0425	5 0.063	0.26	5 0.134	0.0186	0.09	0.5	4.12
W	183.84																	0.05		0.06	0.0372		0.51	0.075
Y	88.906																				0.0017			0.009
Zn	65.38	0.113	0.0011	0.2999	0.075	0.0515	0.111	0.24	0.28	0.19	0.11	0.045	0.0098								0.0163	0.015	0.01	0.068
Zr	91.224	0.015	0.0515	0.239		0.0054																		

EWLA, JUNE 27th, 2018 | PAGE 18



TEMPERATURE AND ELECTRONIC DENSITY

- Local Thermodynamic Equilibrium (LTE) hypothesis
- Lines of neutral species

Boltzmann equation :

Number of analyte atoms in gas phase in the plasma

$$I_{ij} = F(\lambda) \times \frac{hc}{\lambda} \times N_0 \times \frac{g_i A_{ij}}{Z(T_e)} e^{\left(-\frac{E_i}{k}\right)}_{\text{Temperature of the ablation plasma}}$$

Saha equation :

$$\frac{N_{1,X}}{N_{0,X}}N_{e} = \frac{2}{\Lambda^{3}}\frac{Z_{1,X}(T_{e})}{Z_{0,X}(T_{e})}e^{\frac{-E_{X}}{k_{B}T_{e}}}$$

Ceaden Ablated Mass

Correlation between the ablated mass and the physical properties of the matrices



With:

 $L_{th} = 2.\sqrt{D_{th}}.\tau$: Thermal diffusion length

- D_{th} : Thermal diffusivity
- τ : Pulse duration
- n : Real part of the refractive index

J. Picard, "Développement de la LIBS pour l'analyse en ligne de produits uranifères ou plutonifères solides," Dijon, 2015.

Edouard.rollin@cea.fr

Commissariat à l'énergie atomique et aux énergies alternatives DANS Centre de Saclay | 91191 Gif-sur-Yvette Cedex T. +33 (0)1 69 08 71 84 DPC SEARS LANIE

Etablissement public à caractère industriel et commercial RCS Paris B 775 685 019