

# A comparative study using liquid scintillation counting to determine <sup>63</sup>Ni in low and intermediate level radioactive waste

Céline Gautier, Christèle Colin, Cécile Garcia

# ▶ To cite this version:

Céline Gautier, Christèle Colin, Cécile Garcia. A comparative study using liquid scintillation counting to determine  $^{63}$ Ni in low and intermediate level radioactive waste. Journal of Radioanalytical and Nuclear Chemistry, 2016, 308 (1), pp.261-270. 10.1007/s10967-015-4301-4. cea-02383960

# HAL Id: cea-02383960 https://cea.hal.science/cea-02383960

Submitted on 28 Nov 2019

**HAL** is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers. L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.

| 1  | A comparative study using liquid scintillation counting  |
|----|--|
| 2  | to determine <sup>63</sup> Ni in low and intermediate level  |
| 3  | radioactive waste  |
| 4  | Names of the authors: Céline Gautier <sup>1</sup> , Christèle Colin <sup>1</sup> , Cécile Garcia <sup>1‡</sup> |
| 5  | Title: A comparative study using liquid scintillation counting to determine <sup>63</sup> Ni in low            |
| 6  | and intermediate level radioactive waste   |
| 7  | Affiliation(s) and address(es) of the author(s): <sup>1</sup> Operator Support Analyses Laboratory,            |
| 8  | Atomic Energy Commission, CEA Saclay, DEN/DANS/DPC/SEARS/LASE, Building  |
| 9  | 459, PC 171, 91191 Gif-sur-Yvette Cedex, FRANCE  |
| 10 | <sup>#</sup> on leave for AREVA, Demantelement et Services/MSIS Assistance, 91196 Gif-sur-                     |
| 11 | Yvette Cedex, FRANCE   |
| 12 | E-mail address of the corresponding author: celine.gautier@cea.fr  |

| 14 | A comparative study using liquid scintillation counting  |  |  |  |  |
|----|--|--|--|--|--|
| 15 | to determine <sup>63</sup> Ni in low and intermediate level                                      |  |  |  |  |
| 16 | radioactive waste  |  |  |  |  |
| 17 | Céline Gautier <sup>1</sup> , Christèle Colin <sup>1</sup> , Cécile Garcia <sup>1#</sup>         |  |  |  |  |
| 18 | <sup>1</sup> Operator Support Analyses Laboratory, Atomic Energy Commission, CEA Saclay,         |  |  |  |  |
| 19 | DEN/DANS/DPC/SEARS/LASE, Building 459, PC 171, 91191 Gif-sur-Yvette Cedex,                       |  |  |  |  |
| 20 | FRANCE   |  |  |  |  |
| 21 | <sup>#</sup> On leave for AREVA, Demantelement et Services/MSIS Assistance, 91196 Gif-sur-Yvette |  |  |  |  |
| 22 | Cedex, FRANCE  |  |  |  |  |

# 23 Abstract

A comparative study using liquid scintillation counting was performed to measure <sup>63</sup>Ni in 24 25 low and intermediate level radioactive waste. Three dimethylglyoxime (DMG)-based 26 radiochemical procedures (solvent extraction, precipitation, extraction chromatography) 27 were investigated, the solvent extraction method being considered as the reference 28 method. Theoretical speciation calculations enabled to better understand the chemical 29 reactions involved in the three protocols and to optimize them. In comparison to the 30 method based on DMG precipitation, the method based on extraction chromatography 31 allowed to achieve the best results in one single step in term of recovery yield and 32 accuracy for various samples.

# 33 Keywords

<sup>63</sup>Ni, radiochemical analysis, liquid scintillation counting, decommissioning, radioactive
 waste, dimethylglyoxime

# 36 Introduction

37 In France, the National Radioactive Waste Management Agency (ANDRA) is in 38 charge of the long-term management of all radioactive waste. Several repository sites 39 have been built in order to accommodate nuclear waste packages. One is dedicated to the 40 Low and Intermediate Level short-lived Waste. The specifications for 143 radionuclides 41 have been defined by ANDRA which guarantees the safety of the facility [1]. Among this long list,  $^{63}$ Ni has to be declared as soon as its activity concentration is over 1 Bg g<sup>-1</sup> and 42 its maximum acceptance limit has been fixed to 3 x  $10^6$  Bg g<sup>-1</sup> [1]. <sup>63</sup>Ni is produced by 43 44 neutron activation reactions of stable Ni and Cu which are components of various materials used in the nuclear fuel cycle [2]. Consequently, <sup>63</sup>Ni can be present in many 45 46 radioactive materials and waste samples [2-17], such as graphites [6, 7], metals 47 (aluminium, lead, steel) [6-11], concretes [6, 7, 10, 12], ion-exchange resins and 48 charcoals [13], effluents [8, 14-17], sludges [14] and environmental samples [10, 18].

49 <sup>63</sup>Ni is a long-lived radionuclide with a half-life of 98.70 years  $(\pm 24)$  [19]. It is a pure beta emitter with a maximum energy of 66.98 keV [19]. As liquid scintillation counting 50 (LSC) has a high counting efficiency for  $^{63}$ Ni (around 70 %) [2], this detection technique 51 is widely used for <sup>63</sup>Ni determination [2-17]. As a pure beta emitting radionuclide, <sup>63</sup>Ni 52 53 must be isolated from the matrix and the interfering radionuclides (especially <sup>60</sup>Co a 54 major radionuclide which has a similar chemical behavior) through chemical separations 55 prior to any analysis by LSC [2-17]. Consequently, a selective radiochemical method is needed to measure <sup>63</sup>Ni in low and intermediate level radioactive waste [2-18]. Most 56 procedures of <sup>63</sup>Ni purification rely on the complexing agent of dimethylglyoxime 57 58 (DMG) implemented in three different types of methods: solvent extraction, precipitation 59 and extraction chromatography [2-18]. In all cases, the Ni(DMG)<sub>2</sub> complex is favourably 60 formed at basic pH, around 8-9 [2-18]. The recovery yield of the overall radiochemical 61 procedure is generally determined from the measurement of stable Ni by atomic 62 absorption spectroscopy (AAS) [12] or inductively coupled plasma - atomic emission spectroscopy (ICP-AES) [5, 13, 15, 17]. 63

Two or three decades ago, the reference radiochemical method to analyse <sup>63</sup>Ni was 64 based on a liquid-liquid extraction procedure. The Ni(DMG)<sub>2</sub> complex is first extracted 65 66 in an organic solvent [20], commonly chloroform [8, 10, 11, 18, 20] which has a higher 67 Ni extraction capacity [20]. Ni is then back-extracted in aqueous solution, mostly with 68 hydrochloric acid [11, 16, 18]. In France, this extraction method has been standardized in the standard NF M60-317 to determine <sup>63</sup>Ni in radioactive effluents and waste [21]. Ni 69 70 amount is generally less than 1 mg [8, 18, 20] whereas the DMG amount varies from 10 71 mg [20] to 250 mg [8]. By replicating several extractions, this type of separation 72 procedure enabled to achieve satisfactory decontamination factors of Co towards Ni (less 73 than 0.2% of Co was extracted) [8]. In spite of its efficiency, the implementation of this 74 solvent extraction procedure has tended to decrease in the last decades because of the 75 restrictions of chloroform use, notably through the European REACH regulation [22].

76 An alternative method to solvent extraction is the precipitation of the  $Ni(DMG)_2$ complex [4, 9, 12-14]. The French standard NF M60-317 also includes this alternative 77 78 option as a second <sup>63</sup>Ni purification method [21]. When the total activity concentrations of the other radionuclides are 10 times higher in comparison to <sup>63</sup>Ni, this standard 79 80 indicates the necessity to perform a second precipitation step [21]. Higher Ni amount is 81 added (around 2 or 3 mg) [12-14] whereas the DMG amount varies from 50 mg [12, 13] 82 to 200 mg [21] to favour the precipitation of the Ni(DMG)<sub>2</sub> complex, in comparison to the solvent extraction method. Prior to LSC, the precipitate is destroyed to recover <sup>63</sup>Ni in 83 84 solution by using concentrated nitric acid [4, 9, 12, 13] or hydrogen peroxide [14]. The procedure based on Ni(DMG)<sub>2</sub> precipitation has been applied for the measurement of 85 <sup>63</sup>Ni in various radioactive matrices [4], such as metals [9], concretes [12], ion exchange 86 87 resins [13] and sludges [14]. However, the destruction of Ni(DMG)<sub>2</sub> precipitate appears 88 to be a delicate and fastidious step before LSC analysis [21].

To overcome these above problems, the technique of extraction chromatography based on the Eichrom Ni<sup>®</sup> resin has been developed to isolate Ni from the interfering elements [23]. Some authors also prepared in-house Ni resins which relies on the same principle [15, 27]. Indeed, over the past 20 decades, extraction chromatography has become a leading technique for separation and preconcentration of radionuclides in the

94 environmental, biological and nuclear fields [24, 25]. The combination of an organic 95 extractant coated on an inert support delivers the selectivity of solvent extraction with the 96 ease of use of resin based methods. In the case of Ni resin, the DMG extractant is coated 97 on an inert support of acrylic ester based-resin [23]. As relatively high amounts of DMG 98 and Ni are involved (respectively 50 mg and 2 to 3 mg for a 2 mL pre-packed column 99 [23]), on-column precipitation of Ni with DMG occurs on Ni resin [23]. Elimination of 100 the interfering elements is mainly achieved with ammonium citrate during the rinsing 101 step. Then, Ni is generally stripped from the column using nitric acid [23, 26]. In recent 102 years, many radiochemical procedures based on Ni resin have been applied on many 103 nuclear materials [5, 6, 12, 13, 15, 17, 27].

104 DMG is an effective and selective complexing agent of Ni but also of other metal elements, such as Co, Cu, Cd and Pd [28], which can induce interferences for <sup>63</sup>Ni 105 purification. Indeed, the <sup>60</sup>Co activation product is often present in substantial amounts in 106 radioactive materials in comparison to <sup>63</sup>Ni. Correlation factors between <sup>63</sup>Ni and <sup>60</sup>Co 107 highly depend on the types of nuclear plants and samples [29]. In CEA France, the third 108 quartile of <sup>63</sup>Ni/<sup>60</sup>Co ratio has been determined at 0.4 in solid radioactive waste. 109 Consequently, from the literature, it is frequently necessary to complete the purification 110 111 step based on DMG with other separation procedures so as to eliminate Co efficiently. In 112 the French standard NF M60-317, the elimination of Co is achieved with a preliminary 113 liquid-liquid extraction step based on the use of 2-nitroso-1-naphthol [21]. In this 114 standard, it is recommended to implement this Co solvent extraction when the total 115 activity concentrations of the other radionuclides are 10 times higher in comparison to <sup>63</sup>Ni [21]. Furthermore, the presence of <sup>55</sup>Fe, another significant activation product, can 116 117 also hinder the formation of Ni(DMG)<sub>2</sub> complex/precipitate because of its precipitation at 118 basic pH [23, 26]. Organic complexing agents, such as citric acid [6, 12, 21], tartaric acid 119 [9, 21] or oxalic acid [5] are generally introduced to prevent the precipitation of Fe and 120 the other metal elements at basic pH. However, their chelating properties may not be 121 sufficient in case of high Fe amounts, such as in steels [6, 28]. Consequently, it is also highly recommended to remove Fe to achieve accurate <sup>63</sup>Ni measurements. Precipitation 122 with ammonia [12-16, 18] or hydroxide [6, 14] and anion exchange chromatography [4, 123 124 5, 9, 10, 11, 14, 15, 17] have been mainly applied in order to eliminate the interfering

elements such as Co and Fe. Decontamination factors of  $10^5$  or higher can be obtained 125 using a combination of these purification methods [2, 6]. In our group, the reference 126 radiochemical procedure to measure <sup>63</sup>Ni corresponds to the French standard NF M60-127 317 [21]. It consists in combining systematically the Co solvent extraction (based on 2-128 129 nitroso-1-naphthol) with the Ni solvent extraction (based on DMG). This method has 130 been applied to various radioactive waste samples and in particular to graphites [30, 31]. 131 The works reported in the literature generally include several separation steps to analyse <sup>63</sup>Ni [4-18] but do not compare different methods of <sup>63</sup>Ni determination. So, it is not 132 obvious to choose the most effective and simple procedure for the measurement of <sup>63</sup>Ni in 133 134 various low and intermediate level radioactive waste samples.

135 The aim of this work is to propose a reliable radiochemical method to measure <sup>63</sup>Ni in various low and intermediate level radioactive waste with a large range of <sup>63</sup>Ni/<sup>60</sup>Co 136 ratios. For that purpose, the speciation of metal elements (Ni, Co and Fe) is first 137 138 discussed so as to achieve a better understanding of their behaviours in the chemical 139 reactions involved in the different separation steps. Subsequently, three radiochemical 140 procedures based on the use of DMG complexing agent were applied to different 141 radioactive samples (evaporate concentrates, steels, muds embedded in concretes, 142 effluents, ion exchange resins embedded or not in polymers and graphites). Solvent 143 extraction, precipitation and extraction chromatography based on DMG were compared 144 in terms of recovery yield and accuracy to determine the radiochemical method the most 145 selective in one single separation step.

146

### 147 **Experimental**

### 148 Reagents and equipments

All chemicals (nitric acid, hydrochloric acid, hydrofluoric acid, hydroiodic acid, ammonium hydroxide, hydrogen peroxide, citrate ammonium, sodium citrate, tartaric acid, 2-nitroso-1-naphthol, dimethylglyoxime) were of analytical grade. Ultra-pure water

152 (resistivity 18.2 M $\Omega$  cm) was obtained from a Milli-Q purification system (Millipore, 153 France). Anion-exchange resin AG1-X4 (50-100 mesh) was supplied by Bio-Rad 154 Laboratories (France). In-house 0.8 cm x 5 cm columns were prepared with 2 g of AG1-155 X4 resin. Pre-packed Ni<sup>®</sup> cartridges of 2 mL (100-150 mesh) were purchased from 156 Triskem International (France).

All <sup>63</sup>Ni measurements were performed with a Tri-Carb liquid scintillation counter 157 (Perkin-Elmer, France). The instrument was calibrated using a certified <sup>63</sup>Ni standard 158 (NI63ELSB30 having a massic activity of 4 x  $10^4$  Bg g<sup>-1</sup> ± 3.5%) which was purchased 159 from CERCA LEA (France). After the radiochemical procedures, aliquots of 5 mL of the 160 <sup>63</sup>Ni purified samples were mixed with 15 mL Ultima Gold<sup>TM</sup> LLT scintillation cocktail 161 (Perkin-Elmer, France) in 20 mL polyethylene vials (Perkin-Elmer, France). All 162 163 measurements of gamma emitting radionuclides were completed using a high purity 164 germanium detector (Canberra, France) which was calibrated with a multi gamma standard (9ML01ELME20) supplied by CERCA LEA (France). Spectral analysis and 165 166 quantification were carried out with Genie 2000 software. The accuracy of our 167 radiological analyses was previously checked with proficiency tests, in particular with the European Carbowaste project on irradiated graphites [30, 31]. The uncertainties of the 168 169 <sup>63</sup>Ni activities concentrations were calculated according to the standard NF M60-317 [21] 170 by combining the uncertainties associated with the quantities of digested samples, the 171 standards, the recovery yields and the LSC measurements. The overall expanded 172 uncertainties were calculated by using a coverage factor k of 2.

Stable Fe, Co and Ni concentrations were measured using an ICP-AES (Inductively Coupled Plasma - Atomic Emission Spectroscopy) Activa M spectrometer (HORIBA Jobin Yvon, Longjumeau, France). External calibration curves were established from ICP standards (SPEX Certiprep, USA). The accuracy of our elemental analyses was previously checked with proficiency tests, in particular those organized by the French CETAMA and AGLAE committees.

179

# 181 Sample preparation and digestion

182 The different radioactive samples were collected in several French nuclear facilities and laboratories. They consisted of the following types of materials: evaporate 183 184 concentrates, steels, muds embedded in concretes, effluents, ion exchange resins 185 embedded or not in polymers and graphites (denoted from S1 to S8 in this work). They were chosen in order to cover a wide range of <sup>63</sup>Ni/<sup>60</sup>Co ratios (from 0.3 to 16.4). All 186 187 samples were digested using a microwave acid digestion system (Speed Wave, Berghof, 188 Germany), except graphites. The digestion conditions were consistent with the French 189 guide NF M60-323 [32]. The evaporate concentrate sample (denoted as S1), the effluent 190 (denoted as S4) and the ion exchange resins embedded or not in polymers (denoted as S5, 191 S7 and S8) were digested with 15 mL of concentrated HNO<sub>3</sub>. The steel sample (denoted 192 as S2) was digested using aqua regia (5 mL of concentrated HNO<sub>3</sub> and 10 mL of 193 concentrated HCl). The muds embedded in concretes (denoted as S3) were digested using 194 a mixing of 10 mL of concentrated HNO<sub>3</sub> and 5 mL of concentrated HF. The graphite 195 samples were digested by using HI in excess and 5 mL of concentrated H<sub>2</sub>SO<sub>4</sub> so as to 196 decompose graphite as CO<sub>2</sub> by heating. In all cases, 0.2 g up to 5 g of samples were 197 digested. The resulting solutions were transferred to 100 mL volumetric flasks and 198 diluted with ultra-pure water. Aliquots were prepared for the determination of stable 199 metal elements and gamma emitting radionuclides by ICP-AES and gamma spectrometers respectively. As a function of the <sup>63</sup>Ni activity concentrations, 5 mL to 20 200 mL of the digested samples were used for <sup>63</sup>Ni purification. The amount of added Ni 201 202 carrier was then adapted depending on the studied radiochemical methods (from 0.1 mg 203 to 4 mg). As examples, the chemical and radiochemical compositions of two studied samples (S1 and S2) are detailed in Table 1. For S2 steel, <sup>55</sup>Fe activity concentration was 204 determined at 7 x  $10^3$  Bg g<sup>-1</sup> from the French standard NF M60-322 which relies on the 205 206 solvent extraction of the Fe complex formed with cupferron (N-nitroso-N-207 phenylhydroxylamine) in chloroform [33].

# 209 Radiochemical separation

### 210 Method 1 based on the organic extraction of the Ni(DMG)<sub>2</sub> complex

211 Method 1 is based on the organic extraction of the Ni(DMG)<sub>2</sub> complex. It 212 corresponds to one of the separation protocols described in the standard NF M60-317 [21]. This is the reference radiochemical procedure of our group to measure  $^{63}$ Ni in 213 214 radioactive waste and effluents [30]. Aliquots of the digested samples were first weighed 215 in a beaker. 0.1 mg of Ni carrier and 0.5 mg of Co carrier were then added. After the 216 addition of 5 mL concentrated HNO<sub>3</sub>, the solutions were evaporated to dryness and the 217 residues were dissolved in a few mL of HNO<sub>3</sub>. 10 mL of sodium citrate (100 g/L) was 218 then added and the pH of the solutions was adjusted between 3 and 4 with NH<sub>4</sub>OH. 219 Afterwards, 1 mL of concentrated hydrogen peroxide was added to oxidize Co. Then, 4 220 mL of an acetic solution of 2-nitroso-1-naphthol (10 g/L) was introduced. As the reaction 221 between Co and 2-nitroso-1-naphthol proceeds rather slowly [34], the solutions were 222 allowed to stand for about 30 min. 3x10 mL of chloroform were then added to extract the 223 Co-nitrosonaphtol complex in the organic phases whereas Ni remained in the aqueous 224 phases. Thereafter, the pH of the solutions was adjusted to 9 with concentrated NH<sub>4</sub>OH. 2 225 mL of DMG solution (10 g/L in ethanol) was then added. After a waiting period of 1 226 hour, 3x10 mL of chloroform were added to extract the Ni(DMG)<sub>2</sub> complex. Afterwards, 227 the organic phases were washed with 10 mL of 5 % NH<sub>4</sub>OH. Ni was then back-extracted 228 using 3x10 mL of 0.5 M HCl. The obtained solutions were then evaporated to dryness 229 and the residues were dissolved in 10 mL of 0.5 M HCl. Finally, a 5 mL aliquot of the <sup>63</sup>Ni purified samples was prepared for LSC as described above. 230

# 231 Method 2 based on the precipitation of the Ni(DMG)<sub>2</sub> complex

Method 2 is based on the precipitation of the Ni(DMG)<sub>2</sub> complex. It is also described in the standard NF M60-317 [21]. As the total activity concentrations of the other radionuclides are not 10 times higher in comparison to  $^{63}$ Ni in the studied samples, only one precipitation step was implemented from the specifications of this standard [21].

236 Aliquots of the digested samples were first weighed in a beaker. 10 mL of a saturated 237 tartaric solution and 4 mg of Ni carrier were then added. The pH of the solutions was 238 adjusted to 9 with concentrated NH<sub>4</sub>OH. 5 mL of DMG solution (1% in ethanol) was then 239 added. The obtained solutions were heated at 50°C during 30 min so as to favour the 240 precipitation of the Ni(DMG)<sub>2</sub> complex. The precipitates were collected by filtration and 241 washed with water. Thereafter, Ni was recovered by pouring gently 5 M HCl on the 242 precipitate. Afterwards, the solutions were evaporated near to dryness (otherwise 243 insoluble black residues were obtained as indicated in Ref. [26]). DMG was then 244 decomposed with hydrogen peroxide. Thereafter, the obtained solutions were evaporated 245 to dryness and the residues were dissolved in 10 mL of 0.5 M HCl. Finally, a 5 mL aliquot of the <sup>63</sup>Ni purified samples was prepared for LSC as described above. 246

### 247 Method 3 based on the precipitation of the Ni(DMG)<sub>2</sub> complex on Ni resin

248 Method 3 is based on the precipitation of the Ni(DMG)<sub>2</sub> complex on Ni resin [23]. It 249 corresponds to the radiochemical method described by Eichrom Technologies [26] and 250 applied in many reported works [6, 12, 13, 15, 17, 27]. Aliquots of the digested samples 251 were first weighed in a beaker and 2 mg of Ni carrier were then added. After the addition 252 of 5 mL concentrated HCl, the solutions were evaporated to dryness and the residues 253 were dissolved in a few mL of 1 M HCl. 1 mL of 1 M ammonium citrate (that was 254 preliminary adjusted to pH 8-9 with  $NH_4OH$ ) was then added. Afterwards, the pH of the 255 solutions was adjusted between 8 and 9 with NH<sub>4</sub>OH. The samples were then loaded on 256 the prepared Ni columns. The pre-packed Ni columns were preliminary conditioned with 257 20 mL of 0.2 M ammonium citrate that was adjusted to pH 8-9 with NH<sub>4</sub>OH. After the 258 loading of the samples, the Ni columns were rinsed with 20 mL of 0.2 M ammonium 259 citrate (that was adjusted to pH 8-9 with NH<sub>4</sub>OH) to eliminate the interfering elements. 260 Thereafter, Ni was stripped with 5 mL of 3 M HNO<sub>3</sub>. Finally, a 5 mL aliquot of the  $^{63}$ Ni purified samples was prepared for LSC as described above. In those conditions, our 261 group checked that the presence of 3 M HNO<sub>3</sub> and DMG did not induce any quenching 262 effect in LSC by performing the protocol with a <sup>63</sup>Ni standard. Eichrom Technologies 263 264 recommend to eliminate Fe prior to the separation on Ni column [26] but no threshold of

265 Fe concentration is indicated. Given the Fe compositions of the studied samples, an 266 additional purification step was introduced only for the steel sample. From the works of 267 Hou et al. and Rajkovich et al. [6, 23], it was decided to implement a separation on the 268 anion exchange AG1-X4 resin before the purification step on the Ni resin. The AG1-X4 269 resin has indeed a higher loading capacity towards Fe (around 15 mg/g resin) in 270 comparison to TRU resin (5 mg/g resin). In concentrated HCl, Co and Fe are fixed on the 271 anion exchange resin whereas Ni is not retained [4, 5, 6, 13, 15-17]. In the case of steels, 272 2 mg of Ni carrier and 1 mg of Co carrier were added to the aliquots of the digested 273 samples. The solutions were evaporated to dryness and the residues were dissolved in 5 274 mL of 8 M HCl. The samples were then loaded on the AG1-X4 columns which were 275 preliminary conditioned with 25 mL of 8 M HCl. After the loading of the samples, Ni 276 was eluted with 5 mL of 8 M HCl. The resulting solutions were then evaporated to 277 dryness and the residues were treated as described above in the protocol dedicated to the 278 Ni columns.

279

# 280 **Results and discussion**

### 281 Speciation studies

282 A previous work of our group demonstrated the importance of speciation studies in 283 order to achieve a better understanding of the behaviours of the analytes during the 284 different separation steps and to optimize the radiochemical procedures [25]. This 285 approach was also investigated by Rosskopfova et al. so as to determine the Ni species in a method dedicated to <sup>63</sup>Ni purification in nuclear waste [12]. Nonetheless, Rosskopfova 286 287 et al. did not take into account of all the chemical reagents (such as citrate) and the 288 interfering elements (such as Co and Fe) in their speciation calculations [12], which can influence greatly the speciation results. Indeed, as a pure beta emitter, <sup>63</sup>Ni needs to be 289 290 isolated from the matrix and the interfering elements. Consequently, speciation studies 291 were performed by considering the main chemicals and metal elements involved in the

three <sup>63</sup>Ni radiochemical methods studied by our group. For that purpose, speciation 292 293 calculations were made with JChess software (Ecole des Mines ParisTech, France). The 294 database of the software (chess.tdb) was enriched with specific stability constants of Ni, 295 Co and Fe with ammonia, citrate, tartrate and DMG (the stability constants related to 296 hydroxide complexes and precipitates were already included). The relevant stability 297 constants were obtained from Smith and Martell [28]. As citric acid has three acidic 298 functions (the corresponding pKa values are: pKa1 = 3.1, pKa2 = 4.8, pKa3 = 6.4) [28], it 299 was denoted as H<sub>3</sub>Cit. In the same way, tartaric acid was denoted as H<sub>2</sub>Tart (the 300 corresponding pKa values are: pKa1 = 2.7, pKa2 = 3.7) [28]. Simple calculations were 301 first performed by investigating the behaviours of Ni, Co and Fe in NaOH and NH<sub>4</sub>OH. 302 Whatever their quantities, Ni, Co and Fe are predicted to precipitate quantitatively with hydroxide ions by using NaOH. Fe is foreseen to precipitate and can also co-precipitate 303 304 avec Co in NH<sub>4</sub>OH medium, depending on their quantities. With regard to Ni, its predominant species are Ni(NH<sub>3</sub>)<sub>x</sub><sup>2+</sup> complexes by using NH<sub>4</sub>OH: no significant Ni 305 306 precipitation is predicted at basic pH. Those results were in agreement with the 307 experimental studies of Hou et al. [6], which proves the reliability of our JChess 308 calculations.

309 The results obtained from JChess software are presented for the S2 steel sample (its 310 chemical composition is given in Table 1) but similar conclusions can be drawn for the 311 other materials, such as the S1 evaporate concentrate. The theoretical distribution 312 diagrams of Ni(II), Co(II) and Fe(III) species are depicted respectively in Figures 1 to 3 313 for method 1. In the case of method 1, after the treatement of the sample in HNO<sub>3</sub>, sodium citrate was added and pH was adjusted to 3-4 with NH<sub>4</sub>OH. From JChess 314 calculations, the predominant Ni(II) species are  $Ni^{2+}$ ,  $Ni(NO_3)^+$  and Ni-citrate complexes 315 (NiCit<sup>-</sup> and NiH<sub>2</sub>Cit<sup>+</sup>) at pH 3-4. Those results are consistent with the work of Zelenin et 316 317 al. who investigated the interaction of the Ni(II) ion with citric acid in an aqueous 318 solution [35]. It is predicted that Co has the same behaviour as Ni at pH 3-4: the predominant Co(II) species are  $Co^{2+}$ , Co(NO<sub>3</sub>)<sup>+</sup> and Co-citrate complexes (CoCit<sup>-</sup> and 319 320  $CoH_2Cit^+$ ). In those pH conditions, the predominant Fe(III) species are Fe-citrate 321 complexes (FeCit and FeOHCit). For method 1, no precipitation is predicted at pH 3-4. 322 Consequently, after the addition of hydrogen peroxide, Co can be complexed by 2-

323 nitroso-1-naphthol and further extracted by chloroform [34], which induces an efficient 324 elimination of Co. From literature [34], Fe can also be partly extracted. In this step, Ni 325 which was not complexed by 2-nitroso-1-naphthol remained in the aqueous phase. The 326 pH of the aqueous solution was then adjusted to 9 with ammonia. The speciation studies indicate that the predominant Ni(II) species are Ni(NH<sub>3</sub>)<sub>x</sub><sup>2+</sup> complexes at pH 9, which 327 328 prevents Ni from precipitating at basic pH. This is not the case for Fe which might start 329 precipitating as  $Fe_2O_3$  hematite when the pH is higher than 6. After the pH adjustment at 330 9, DMG was added. It can be noted that no stability constant was found for Fe with 331 DMG: no stable complex might be formed [28]. On the contrary, the predominant Ni(II) 332 species is the Ni(DMG)<sub>2</sub> complex at pH 9, which is in agreement with Dyrssen et al. [36]. 333 Consequently, only Ni can be extracted in the organic phase after the addition of 334 chloroform, which enables a selective purification of Ni for method 1. However, the 335 presence of Fe precipitate might hinder the Ni extraction and induce a slight decrease of 336 Ni recovery yield. Finally, Ni can be back-extracted at low pH in HCl medium for LSC 337 analysis.

338 Concerning method 2, after the addition of tartaric acid and the pH adjustment with ammonia, the same Ni(II) species as above might be observed: the Ni(NH<sub>3</sub>) $_{x}^{2+}$  complexes 339 are predicted to be predominant at pH 8-9. In contrast, despite the use of tartaric acid, Co 340 341 and Fe are foreseen to precipitate as  $CoFe_2O_4$  and  $Fe_2O_3$  at basic pH. The complexing 342 properties of this chelating agent are not sufficient to prevent Co and Fe precipitation in 343 the case of steels (the formation constant of  $Fe_2O_3$  hematite is much greater than the one 344 of FeTart<sup>+</sup> [28]). Afterwards, DMG was introduced in the aqueous solution. Since high 345 amounts of Ni and DMG were used for method 2, the Ni(DMG)<sub>2</sub> precipitate is predicted 346 to be formed. As Co and Fe are present as precipitates in the former step, a co-347 precipitation with Ni(DMG)<sub>2</sub> might occur. From speciation calculations, the main 348 interfering elements, Co and Fe, are predicted to be not eliminated efficiently with 349 method 2 in the case of a steel sample. Similar results were obtained for the other 350 samples depending on Co and Fe quantities. It can be noticed that this point was not 351 precisely specified in the French standard [21]. After the DMG precipitation step, the precipitate can be dissolved at low pH in concentrated HCl to recover Ni in solution. 352

353 The distribution diagrams of Ni(II), Co(II) and Fe(III) species related to method 3 are 354 very close to those obtained for method 1 (see Figures 1 to 3). After the addition of 355 ammonium citrate and the pH adjustment with ammonia, the predominant Ni(II) species are  $Ni(NH_3)_x^{2+}$  complexes at pH 8-9. As the used quantity of citrate is quite low in 356 357 method 3, at pH 8-9, Co and Fe exist as precipitate of  $CoFe_2O_4$  and  $Fe_2O_3$  in the case of 358 steel sample. Similar conclusions were obtained for the other samples. Consequently, 359 cautiousness has to be taken towards method 3 when Co or Fe are present in the analysed 360 aliquots, which is in agreement with Eichrom Technologies' recommendations [26]. 361 After the pH adjustment, the sample was loaded on the Ni resin which was preliminary 362 conditioned with ammonium citrate at pH 8-9. Since high amounts of Ni and DMG are 363 used for method 3, the Ni(DMG)<sub>2</sub> precipitate is predicted to be formed on the Ni column. After the rinsing step of the Ni column, concentrated HNO<sub>3</sub> was introduced. At pH lower 364 than 1. the predominant Ni(II) species is predicted to be  $Ni^{2+}$  which allowed the Ni 365 366 stripping from the column and its further LSC measurement.

367 From the speciation studies, it can be inferred that the behaviours of Ni, Co and Fe 368 strongly depend on the considered purification methods. For method 1, no significant 369 reaction was predicted to hinder the selective isolation of Ni from Co and Fe. On the 370 contrary, Co and Fe might precipitate quantitatively during the different steps of method 371 2 and 3 as a function of their concentrations, but these interfering reactions are predicted 372 to be more pronounced for method 2. For a steel sample, according to JChess 373 calculations, the quantity of the complexing agent (citrate or tartrate) should be increased 374 by a 100 fold factor to prevent any Co and Fe precipitation, which is not convenient in 375 practice. As a consequence, it might be quite difficult to avoid the presence of interferences for <sup>63</sup>Ni purification by applying only method 2, in case of high amounts of 376 377 Co and Fe in the samples.

378

Applications of the three DMG-based radiochemical methods to real
 nuclear waste samples

382 From the literature, it is not obvious to choose the best selective radiochemical 383 procedure based on DMG (solvent extraction, precipitation and extraction 384 chromatography) and to determine if a single separation step can be implemented for 385 nuclear waste samples. For instance, for method 3 based on Ni column, Rajkovich et al. 386 [23] reported a decontamination factor of 100000 for Co. On the other hand, Hou et al. 387 [5] obtained a decontamination factor of 2000 for Co whereas Warwick et al. [15] 388 indicated the presence of 5 % of Co in the Ni purified fractions. To answer this question, 389 the three radiochemical methods based on the use of DMG complexing agent were 390 applied to different types of radioactive waste samples (evaporate concentrates, steels, 391 muds embedded in concretes, effluents, ion exchange resins embedded or not in polymers 392 and graphites). The samples denoted from S1 to S8 were chosen so as to cover a large range of <sup>63</sup>Ni/<sup>60</sup>Co ratios from 0.3 to 16.4. In those conditions, less than 2 % of Co must 393 394 be present in the Ni purified fractions so that Co does not induce any interference in LSC 395 analysis.

The results obtained for the three DMG-based radiochemical methods are given in Table 2 in terms of  ${}^{60}$ Co and  ${}^{63}$ Ni activity concentrations. It can be noted that a wide range of  ${}^{63}$ Ni activity concentrations were measured from 1 to 2 x 10<sup>6</sup> Bq g<sup>-1</sup> in the eight samples. For all the studied radioactive waste, the  ${}^{63}$ Ni values were lower than the maximum acceptance limit fixed by ANDRA at 3 x 10<sup>6</sup> Bq g<sup>-1</sup> [1].

The three DMG-based radiochemical procedures were compared in terms of Ni recovery yields in Figure 4. The Ni recovery yields were satisfactory whatever the analysed samples and the radiochemical methods. Those results are in agreement with the values reported in the literature [4-18]. For method 1 based on solvent extraction, it can be noticed that the Ni recovery yields varied from 69 % to 97 %. The lower value of Ni recovery yield was obtained for the steel sample purified with method 1. This could be explained by the Fe precipitation which might hinder slightly the extraction of Ni(DMG)<sub>2</sub>

408 complex in the organic phase at pH 9, in agreement with the theoretical speciation
409 calculations. Whatever the sample, the values of Ni recovery yields were higher than 95
410 % and 90 % respectively for methods 2 and 3.

411 The three DMG-based radiochemical methods were then compared in terms of 412 normalized errors in Figure 5, the method 1 based on solvent extraction being considered 413 as the reference method. The formula used for calculating the normalized error E-normal 414 (denoted as  $E_n$ ) is shown in Equation 1 [37]. When  $E_n$  is lower than 1, the performances 415 of the studied method are judged as satisfactory.

416 
$$E_{n} = \frac{\left|x_{i} - X_{method1}\right|}{\sqrt{\left(U_{i}^{2} + U_{method1}^{2}\right)}} \qquad Eq. (1)$$

417 Where

418  $x_i$  is the <sup>63</sup>Ni measurement result obtained with method i (i=2 to 3)

419  $X_{method 1}$  is the <sup>63</sup>Ni reference measurement result obtained with method 1

420  $U_i$  is the expanded uncertainty of  $x_i$ 

421  $U_{method 1}$  is the expanded uncertainty of  $X_{method 1}$ .

422 For method 2 based on DMG precipitation, the E<sub>n</sub> values were higher than 1 for the 423 majority of the studied samples (up to 17), which indicated that the performances of 424 method 2 were unsatisfactory. For instance, the En value related to the S2 steel was 425 around 2. In agreement with the theoretical speciation calculations, those unacceptable 426 results might be related to the presence of Fe and Co in the Ni purified fractions due to 427 their precipitations at basic pH. This hypothesis was confirmed by gamma and ICP-AES measurements. Around 2 % of Fe (thus  ${}^{55}$ Fe) and 2 % of  ${}^{60}$ Co were respectively 428 determined in the Ni fractions, which induced interferences for <sup>63</sup>Ni characterization by 429 LSC, given the  ${}^{63}$ Ni/( ${}^{60}$ Co+ ${}^{55}$ Fe) ratio of 0.2 and  ${}^{63}$ Ni/ ${}^{60}$ Co ratio of 0.4 in the S2 steel. As 430 a conclusion, method 2 implemented with a single precipitation step was not selective 431 enough to ensure accurate <sup>63</sup>Ni measurements in nuclear waste having a large range of 432 433 activity concentrations of interfering elements. Besides, the French standard NF M60-317 434 prescribes that one DMG-based precipitation step should be sufficient when the total

435 activity concentrations of the other radionuclides are not 10 times higher in comparison in comparison to <sup>63</sup>Ni [21], which is not in agreement with our experimental results. This 436 437 standard published in 2001 [21] might be no more adapted to the nuclear waste produced 438 in the past decade from decommissioning operations of various radioactive facilities. 439 Consequently, the French standard NF M60-317 might be revised to take into account of 440 the variety of the present nuclear waste samples. Co and Fe elimination should be 441 specified by performing a second systematic purification step which can be a second 442 DMG-precipitation or a preliminary separation on an anion exchange resin in HCl 443 medium.

444 For method 3 based on Ni column, the  $E_n$  values were lower than 1 whatever the 445 studied samples, which demonstrated that the performances of method 3 were 446 satisfactory. Method 3 proved to be very efficient towards a large variety of radioactive waste. In conclusion, method 3 implemented in only one single Ni column step was 447 selective enough to obtain accurate <sup>63</sup>Ni analyses in nuclear waste with various <sup>63</sup>Ni/<sup>60</sup>Co 448 ratios, except samples with high Fe contents such as steels. In the case of steels, a 449 450 preliminary separation on an anion exchange resin was implemented prior to the Ni resin, which provided reliable  $^{63}$ Ni measurements (the E<sub>n</sub> value was 0.6). Indeed, less than 2 % 451 of Fe (thus <sup>55</sup>Fe) and 2 % of <sup>60</sup>Co were respectively measured in the Ni fractions. 452 Consequently, from these studies, method 3 was deduced to be more selective than 453 454 method 2 in the case of a single separation step.

455 A systematic separation on an anion exchange resin might be proposed prior to the Ni 456 resin in order to obtain a simple and reliable radiochemical method that can be applied to all nuclear matrices whatever their <sup>63</sup>Ni/<sup>60</sup>Co ratios and their Fe amounts. To check this 457 458 assumption, the S3 sample (corresponding to muds embedded in concrete) was analysed 459 with and without performing the AG1-X4 separation before the purification of the Ni column. The difference between the two  $^{63}$ Ni values was less than 1 % and the values of 460 461 Ni recovery yields were very similar (higher than 90 %). Furthermore, a blank separation was implemented so as to determine the <sup>63</sup>Ni detection limit of the overall method 3. The 462 <sup>63</sup>Ni detection limit was found to be 0.2 Bq  $g^{-1}$  for an effluent sample, which is 463 compatible with the  $^{63}$ Ni declaration threshold fixed by ANDRA at 1 Bq g<sup>-1</sup>. As a 464

465 conclusion, method 3 based on the AG1-X4 and Ni resins was selected to determine <sup>63</sup>Ni 466 accurately in low and intermediate radioactive waste whatever their chemical and 467 radiological compositions. In the future, this method which prevents the use of 468 chloroform might be included in the revised version of the French standard NF M60-317.

469

### 470 **Conclusions**

<sup>63</sup>Ni is a major activation product which has to be characterized in low and 471 intermediate level radioactive waste. As a pure beta emitter, this radionuclide must be 472 isolated from the matrix and the interfering elements (mainly <sup>60</sup>Co and <sup>55</sup>Fe) through 473 474 separation procedures prior to LSC measurement. A comparative study using LSC was performed to measure <sup>63</sup>Ni in various radioactive waste samples with a large range of 475 <sup>63</sup>Ni/<sup>60</sup>Co ratios. Three dimethylglyoxime (DMG)-based radiochemical procedures 476 477 (solvent extraction, precipitation, extraction chromatography) were investigated, the 478 solvent extraction method being considered as the reference method. Theoretical 479 speciation calculations enabled to better understand the chemical reactions involved in 480 the different steps of the three protocols and to optimize them. The presence of 481 significant cobalt and iron quantities was predicted to generate interferences for the 482 method based on DMG precipitation, in agreement with the experimental results. The 483 three DMG-based methods were compared in terms of recovery yield and accuracy. In 484 comparison to the method based on DMG precipitation, the method based on nickel 485 extraction chromatography resin allowed to achieve the most reliable results in one single step for the majority of radioactive waste. For the characterization of <sup>63</sup>Ni in all nuclear 486 487 waste samples, the combination of a separation on an anion exchange resin and a 488 purification on a Ni resin was proved to be selective enough to ensure accurate results. 489 The methodology implemented in this work (based on speciation calculations and 490 experimental results) can be applied to future developments of radiochemical procedures, such as <sup>55</sup>Fe determination in nuclear waste. 491

### 492 Acknowledgements

493 The authors thank the organizing committees of 17<sup>th</sup> RadChem and ERA12 conferences 494 for allowing the oral presentations of this work.

# 495 **References**

496 1. ANDRA, National Radioactive Waste Management Agency (2014)
497 ACO.SP.ASRE.99.0002D ANDRA specifications. Accessed 5 June 2015

498 2. Hou X, Roos P (2008) Critical comparison of radiometric and mass spectrometric
499 methods for the determination of radionuclides in environmental, biological and nuclear
500 waste samples. Anal Chim Acta 608:105-139

- 3. Hoeppener-Kramar U, Pimpl M, Willmann F (1997) Application of procedures for low
  level radionuclide analysis in environmental monitoring for the purpose of clearance
  measurements of materials from decommissioning of nuclear facilities. J Radioanal Nucl
  Chem 226:99-103
- 4. Lee CH, Lee MH, Ha YK, Song KS (2011) Systematic radiochemical separation for
  the determination of <sup>99</sup>Tc, <sup>90</sup>Sr, <sup>94</sup>Nb, <sup>55</sup>Fe and <sup>59,63</sup>Ni in low and intermediate radioactive
  waste samples. J Radioanal Nucl Chem 288:319-325

508 5. Lee CH, Choi KS, Song BC, Ha YK, Song K (2013) Rapid separation of nickel for
 <sup>59</sup>Ni and <sup>63</sup>Ni activity measurement in radioactive waste samples. J Radioanal Nucl Chem
 510 298:1221-1226

6. Hou X, Østergaard LF, Nielsen SP (2005) Determination of <sup>63</sup>Ni and <sup>55</sup>Fe in nuclear
waste samples using radiochemical separation and liquid scintillation counting. Anal
Chim Acta 535:297-307

514 7. Hou X (2007) Radiochemical analysis of radionuclides difficult to measure for waste
515 characterization in decommissioning of nuclear facilities. J Radioanal Nucl Chem
516 273:43-48

- 517 8. Poletiko C (1988) Determination of nickel-63. Environ Int 14:387-390
- 518 9. Numajiri M, Oki Y, Suzuki T, Miura T, Taira M, Kanda Y, Kondo K (1994)
- 519 Estimation of nickel-63 in steel and copper activated at high-energy accelerator facilities.
  520 Appl Radiat Isot 45:509-514
- 521 10. Scheuerer C, Schupfner R, Schottelkopf H (1995) A very sensitive LSC procedure to
  522 determine Ni-63 in environmental samples, steel and concrete. J Radioanal Nucl Chem
  523 193:127-131
- 524 11. Shizuma K, Iwatani K, Hasai H, Oka T, Hoshi M, Shibata S, Imamura M, Shibata T
- 525 (1997) Identification of <sup>63</sup>Ni and <sup>60</sup>Co produced in a steel sample by thermal neutrons
- 526 from the Hiroshima atomic bomb. Nucl Instrum Methods Phys Res Sect A 384:375-379
- 12. Rosskopfova O, Galambo M, Rajec P (2011) Determination of <sup>63</sup>Ni in low level solid
   radioactive waste. J Radioanal Nucl Chem 289:251-256
- 529 13. Taddei MHT, Macacini JF, Vicente R, Marumo JT, Sakata SK, Terremoto LAA
- 530 (2013) Determination of <sup>63</sup>Ni and <sup>59</sup>Ni in spent ion-exchange resin and activated charcoal

from the IEA-R1 nuclear research reactor. Appl Radiat Isot 77:50-55

- 532 14. Kaye JH, Strebin RS, Nevissi AE (1994) Measurement of <sup>63</sup>Ni in highly radioactive
  533 Hanford waste by liquid scintillation couting. J Radioanal Nucl Chem 180:197-200
- 15. Warwick PE, Croudace IW (2006) Isolation and quantification of <sup>55</sup>Fe and <sup>63</sup>Ni in
  reactor effluents using extraction chromatography and liquid scintillation analysis. Anal
  Chim Acta 567:277-285
- 537 16. Jordan N, Michel H, Barci-Funel G, Barci V (2008) Radiochemical procedure and
  538 quantitative determination of the activation product, <sup>63</sup>Ni, in environmental soft water
  539 samples with high Ca and Mg phosphate concentration. J Radioanal Nucl Chem 275:253540 256

- 541 17. Remenec B, Dulanska S, Matel L (2013) Determination of difficult to measure
- 542 radionuclides in primary circuit facilities of NPP V1 Jaslovske Bohunice. J Radioanal
- 543 Nucl Chem 298:1879-1884
- 544 18. Holm E, Rots P, Skwarzec B (1992) Radioanalytical Studies of Fallout <sup>63</sup>Ni. Appl
  545 Radiat Isot 43:371-376
- 546 19. Laboratoire National Henri Becquerel (2005) Table de radionucléides <sup>63</sup>Ni.
- 547 20. Yonezawa C, Sagawa T, Hoshi M, Tachikama (1983) Rapid determination of specific
  548 activity of nickel-63. J Radioanal Nucl Chem 78:7-14
- 549 21. AFNOR Standard NF M60-317 (2001) Nuclear energy Nuclear fuel technology -
- 550 Waste Determination of nickel 63 in effluents and waste by liquid scintillation after a
- 551 preliminary chemical extraction. Association Française de Normalisation, Paris, France
- 552 22. European Chemicals Agency (2012) Guidance for the implementation of REACH
  553 (Registration, Evaluation, Authorisation and Restriction of Chemicals). Accessed 5 June
  554 2015
- 23. Rajkovich S, Cahill D, Peedin L, Wheland S, Lardy M (1996) 2 case studies using
  Eichrom's Nickel resin: a nuclear power plant and a commercial laboratory. Eichrom
  Cincinnati Users' Seminar, USA. Accessed 5 June 2015

558 24. Horwitz EP, Dietz ML, Chiarizia R, Diamond H, Maxwell SL, Nelson MR (1995)
559 Separation and preconcentration of actinides by extraction chromatography using a
560 supported liquid anion exchanger: application to the characterization of high-level
561 nuclear waste solutions. Anal Chim Acta 310:63-78

- 562 25. Gautier C, Coppo M, Caussignac C, Fichet P, Goutelard F (2013) Zr and U
  563 determination at trace level in simulated deep groundwater by Q ICP-MS using extraction
  564 chromatography. Talanta 106:1-7
- 565 26. Eichrom Technologies, Inc. (2003) Analytical procedures NIW01, nickel 63/59 in
  566 water, Feb 25

567 27. Fisera O, Sebesta F (2010) Determination of <sup>59</sup>Ni in radioactive waste. J Radioanal
568 Nucl Chem 286:713-717

569 28. Smith RM, Martell AE (1973) In Critical Stability Constants, Plenum Press

570 29. International Atomic Energy Agency (2009) Determination and Use of Scaling
571 Factors for Waste Characterization in Nuclear Power Plants. AIEA, Nuclear Energy
572 Series NW-T-1.18

573 30. Fréchou C, Degros JP (2006) Radiological inventory of irradiated graphite samples. J
574 Radioanal Nucl Chem 273:677-681

575 31. Banford AW, Eccles H, Graves MJ, von Lensa W, Norris S (2008) Carbowaste - An

576 integrated approach to irradiated graphite. Nucl Future 4:1-5

- 577 32. AFNOR Standard NF M60-323 (2011) Nuclear energy Nuclear fuel cycle
  578 technology Waste Guide for pre-analysis dissolution of effluents, waste and
  579 embedding matrices. Association Française de Normalisation, Paris, France
- 33. AFNOR Standard NF M60-322 (2005) Nuclear energy Nuclear fuel cycle
   technology Waste Determination of iron 55 activity in effluents and waste by liquid
   scintillation after prior chemical separation
- 583 34. Marczenko Z, Balcerzak M (2000) In: Separation, Preconcentration and
  584 Spectrophotometry in Inorganic Analysis, Elsevier, Amsterdam, Netherlands
- 585 35. Zelenin O.Y (2007) Interaction of the Ni<sup>2+</sup> Ion with Citric Acid in an Aqueous
  586 Solution. Russ. J. Coord. Chem. 33:346-350
- 587 36. Dyrssen D, Krašovek F, Sillén LG (1959) On the Complex Formation of Nickel with
  588 Dimethylglyoxime. Acta Chem Scand 13:50-59
- 589 37. Standard NF EN ISO/CEI 17043 (2010) General requirements for proficiency testing.
  590 Association Française de Normalisation, Paris, France

591

|                             | Activity<br>concentration<br>( <sup>60</sup> Co) (Bq g <sup>-1</sup> ) | Activity<br>concentration<br>( <sup>137</sup> Cs) (Bq g <sup>-1</sup> ) | Fe<br>(mg g <sup>-1</sup> ) | Co<br>(mg g <sup>-1</sup> ) | Ni<br>(mg g <sup>-1</sup> ) |
|-----------------------------|--|---|-----------------------------|-----------------------------|-----------------------------|
| Evaporate<br>concentrate S1 | $8.64 \ge 10^2 \pm 10\%$   | $1.66 \ge 10^3 \pm 6\%$   | $0.4\pm10\%$                | < 0.05                      | $0.4 \pm 10\%$              |
| Steel S2                    | $1.10 \ge 10^4 \pm 4\%$  | $9.88 \ge 10^2 \pm 10\%$  | $970 \pm 5\%$               | $0.2 \pm 10\%$              | $6 \pm 10\%$                |

# 592 **Table 1** Chemical and radiochemical compositions of studied nuclear waste samples

593

**Table 2** Comparison of the three studied DMG-based radiochemical methods in terms of  $^{60}$ Co and  $^{63}$ Ni activity concentrations (n/a\*: not available, <sup>\*\*</sup>an anion exchange resin is implemented prior to Ni column)

|  | Activity   | Activity                                       | Datio                                     |                                 |  |
|--|--|--|---|---------------------------------|--|
|  | concentration<br>( <sup>60</sup> Co) (Bq g <sup>-1</sup> ) | Method 1 based<br>on DMG<br>solvent extraction | Method 2 based<br>on DMG<br>precipitation | Method 3 based<br>on Ni column  | <sup>63</sup> Ni/ <sup>60</sup> Co<br>for method 1 |
| Evaporate concentrate S1                         | $8.64 \ x \ 10^2 \pm 10\%$                                 | $2.76 \text{ x } 10^2 \pm 5\%$                 | $2.50 \ge 10^2 \pm 5\%$                   | $2.88 \times 10^2 \pm 5\%$      | 0.3  |
| Steel S2   | $1.10 \text{ x } 10^4 \pm 4\%$                             | $3.99 \times 10^3 \pm 5\%$                     | $3.37 \times 10^3 \pm 5\%$                | $3.82 \times 10^3 \pm 5\%^{**}$ | 0.4  |
| Muds embedded<br>in concrete S3                  | $1.38 \ge 10^4 \pm 4\%$                                    | $8.94 \text{ x } 10^3 \pm 5\%$                 | n/a*                                      | $9.35 \ge 10^3 \pm 5\%$         | 0.7  |
| Effluent S4                                      | $2.10\pm10\%$  | $1.39\pm9\%$                                   | $1.18 \ge 10^1 \pm 5\%$                   | $1.45 \pm 9\%$                  | 0.7  |
| Ion exchange<br>resins S5                        | $6.65 \ge 10^5 \pm 4\%$                                    | $1.81 \ge 10^6 \pm 4\%$                        | $1.76 \ge 10^6 \pm 4\%$                   | $1.73 \ge 10^6 \pm 4\%$         | 2.7  |
| Graphite S6                                      | $4.05 \text{ x } 10^3 \pm 4\%$                             | $2.78 \times 10^4 \pm 4\%$                     | n/a*                                      | $2.90 \text{ x } 10^4 \pm 4\%$  | 6.9  |
| Ion exchange<br>resins embedded<br>in polymer S7 | $2.85 \times 10^4 \pm 4\%$                                 | $3.48 \times 10^5 \pm 4\%$                     | $3.26 \ge 10^5 \pm 4\%$                   | $3.50 \ge 10^5 \pm 4\%$         | 12.3   |
| Ion exchange<br>resins S8                        | $4.67 \times 10^3 \pm 4\%$                                 | $7.66 \ge 10^4 \pm 4\%$                        | $6.57 \ge 10^4 \pm 4\%$                   | $7.95 \text{ x } 10^4 \pm 4\%$  | 16.4   |

597

- 599 Fig. 1 Theoretical distribution diagram of Ni(II) species for a steel (using JChess600 software)
- Fig. 2 Theoretical distribution diagram of Co(II) species for a steel (using JChesssoftware)
- Fig. 3 Theoretical distribution diagram of Fe(III) species for a steel (using JChesssoftware)
- Fig. 4 Ni recovery yields (%) obtained for the three studied DMG-based radiochemicalmethods
- 607 Fig. 5 Comparison of the studied DMG-based radiochemical methods in terms of 608 normalized errors  $E_n$



610 Fig. 1 Theoretical distribution diagram of Ni(II) species for a steel (using JChess611 software)













**Fig. 5** Comparison of the studied DMG-based radiochemical methods in terms of 624 normalized errors  $E_n$  (n/a: not available)

