

# Validation of a Monte Carlo prediction model for portal images using PENELOPE

Delphine Lazaro-Ponthus, Juan-Carlos García Hernández, C. Driol, Bénédicte Poumarède, Samuel Legoupil

## ► To cite this version:

Delphine Lazaro-Ponthus, Juan-Carlos García Hernández, C. Driol, Bénédicte Poumarède, Samuel Legoupil. Validation of a Monte Carlo prediction model for portal images using PENELOPE. 11th international workshop on portal imaging (EPI2kX), Jun 2010, Leuven, Belgium. 88, pp.31 - 46, 2010. cea-02558112

# HAL Id: cea-02558112 https://cea.hal.science/cea-02558112

Submitted on 30 Apr 2020  $\,$ 

**HAL** is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers. L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.

## Validation of a Monte Carlo prediction model for portal images using PENELOPE

D. Lazaro-Ponthus, J.-C. Garcia-Hernandez, C. Driol, B. Poumarède and S. Legoupil

<sup>1</sup> CEA, LIST, F-91191 Gif-sur-Yvette, France.

Contact : delphine.lazaro@cea.fr



- Online verification of dose delivery during radiotherapy treatments has become essential in order to ensure that the dose planned by the treatment planning system (TPS) is delivered as accurately as possible and to detect possible deviations.
- One strategy using EPIDs for dosimetric verification consists in comparing predicted dose images with acquired portal images before and/or during treatment [1]. • Our goal is twofold:
- •demonstrate that the Monte-Carlo (MC) simulation code PENELOPE [2] can be used to compute reliably predicted dose images,

•develop and optimize a MC model of an EPID and validate it against measurements.

#### **Measurements**

#### Description of the experimental setup

- Saturne 43 Linac (GEMS)
- 12 MV photons, 200 UM (1,7 Gy/min at 100 cm),
- field sizes: 4x4, 10x10 and 15x15 cm<sup>2</sup>.
- Fluoroscopic EPID (Lynx 2D, Fimel, France)
- fluorescent screen (1 mm copper coated with a GOS fluorescent layer) viewed by a CCD camera via a mirror tilted at an angle of 45°,
- active area: 30x30 cm<sup>2</sup>,

Acquisitions

• pixel size: 0.5x0.5 mm<sup>2</sup> (600x600 pixels).



Fig. 1: a) Experimental setup and b) fluoroscopic EPID

- Source-to-isocenter = 100 cm; source detector distance (SDD) = 150 cm,
- Phantom: 30x30x30 cm<sup>3</sup> water tank (entrance face at 90 cm from the source),
- Each acquired image formed by averaging 15 images of 10 s each,
- Acquisition of open-field portal images with and without the phantom in the beam ("in-air" images), for the three fields sizes (4x4, 10x10 and 15x15 cm<sup>2</sup>)

### Monte Carlo simulations

#### Computation of Monte-Carlo portal images in two steps

- Step 1: computation of a phase-space file (PSF) storing data about particles exiting the linac (150 million), using a MC commissioned model [3]
- **Step 2**: computation of a 150x150 portal image (pixel size: 2 mm) using the PSF as input data by scoring the energy deposit in the GOS layer
  - photon and electron/positron cut-offs set to 0.01 MeV and 0.70 MeV.

#### Investigated EPID models and simulated configurations

 Optimal EPID model determined by simulating 10x10 cm<sup>2</sup> open fields for different models



### Results

#### Determination of the optimal EPID model

Determination of the backscatter compartment thickness:

- calculation of the experimental ratio R defined by:  $R_{exp} = I_{exp}^{10x10 \text{ cm}^2} / I_{exp}^{15x15 \text{ cm}^2}$  (with I: average value in a 6x6 cm<sup>2</sup> central area of the image),
- variation of the backscatter compartment tickness until the simulated R value matches the experimental R value.





- a backscatter compartment must be included in the EPID model,
- the optimal model was obtained with a
- 8 cm water backscatter compartment,
- the EPID signal is not significantly affected by the mirror and the shielding (profiles for models 1 and 2
- almost identical), • the EPID signal is increased by optical photons backscattering within the
- EPID structure.



- γ-index < 1 for 90% of the pixels for large field sizes and <1 for 96% of the pixels for fields of size 10×10 cm<sup>2</sup> or less,
- results a little bit degraded outside the field for large field sizes because the cross-talk effect is not corrected in the measured portal images.

#### **Conclusions and perspectives**

- Portal images can be computed accurately, with and without phantom in the beam, by implementing with the MC code PENELOPE a 3-layer EPID model including a backscatter compartment, whose thickness must be adjusted carefully,
- This model will be used to compute MC predicted portal images for pre-treatment or online verification in radiotherapy.

#### References

- [1] W. van Elmpt et al., 2008 Radioth. Oncol. 88 289-309
- [2] J. Baro et al., 1995 Nucl. Instr. and Meth. B 100 31-46
- [3] B. Habib et al., 2009 Physica Medica in press

