

Measurement of the absolute gamma-ray emission intensities from the decay of ^{103}Pd

Jonathan Riffaud, Philippe Cassette, Margot Corbel, Marie-Christine Lepy, Valérie Lourenço, Mark Kellett

► To cite this version:

Jonathan Riffaud, Philippe Cassette, Margot Corbel, Marie-Christine Lepy, Valérie Lourenço, et al.. Measurement of the absolute gamma-ray emission intensities from the decay of ¹⁰³Pd. IRCM 2019 - 22nd International Conference on Radionuclide Metrology and its Applications, May 2019, Salamanca, Spain. ICRM technical series on radionuclide metrology, ISSN: 2522-4328, 167, pp.111, 2021, Proceedings of the 22nd International Conference on Radionuclide Metrology and its Applications (ICRM 2019). cea-03044492

HAL Id: cea-03044492 https://cea.hal.science/cea-03044492

Submitted on 8 Dec 2020 $\,$

HAL is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers. L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.





Jonathan RIFFAUD, Philippe CASSETTE, Margot CORBEL, Mark A. KELLETT, Marie-Christine LÉPY*, Valérie LOURENÇO CEA, LIST, Laboratoire National Henri Becquerel, F-91191 Gif-sur-Yvette, France.



LABORATOIRE NATIONAL

DE MÉTROLOGIE ET D'ESSAIS

MEASUREMENT OF THE ABSOLUTE GAMMA-RAY EMISSION INTENSITIES FROM THE DECAY OF ¹⁰³Pd



Introduction	Sources preparation
Palladium-103 is a radioisotope of interest in medical applications which is used in brachytherapy implants for the treatment of prostate cancer.	Initial solution : Palladium chloride in ammonium hydroxyde Mass activity ~ 37 GBq.g ⁻¹
It decays through electron capture to excited levels of ¹⁰³ Rh, and especially (99.959%) to the 39.748 keV metastable state.	6 point sources for gamma- and X-ray spectrometry
The radioactive equilibrium between 103 Pd (half-life = 16.991 (19) d) and 103m Rh	Dilution to 10 MBq.g ⁻¹ using 3 mol·L ⁻¹ HCI

(half-life = 56.114 (12) min) is reached within about 9 hours.





Energies are quoted in keV

with 10 µg.g⁻¹ of stable ruthenium to ensure chemical stability of ^{103m}Rh in equilibrium with ¹⁰³Pd.

• 6 volume sources for liquid scintillation (LS)

New dilution by a factor of 20.



Each source contained about 5 kBq of 103 Pd in Ultima Gold scintillator, and about 100 µL of water was added for stabilization purposes.

The impurity checking on the initial solution did not reveal any impurity and the detection limit was less than $2 \cdot 10^{-4}$ Bq·Bq⁻¹. A complementary measurement was performed one month after the activity measurement, to benefit from the ¹⁰³Pd decay: Rhodium-101 (half-life = 3.30 (10) a) was detected with an activity relative to that of ¹⁰³Pd equal to 10^{-5} Bq·Bq⁻¹.

Activity measurement

The palladium chloride solution was standardized by liquid scintillation, using the Triple-to-Double Coincidence Ratio (TDCR) method.

- The LS sources were measured using the RCTD1 counter of LNHB.
- Each source was measured 10 times for 1 minute, allowing relative counting uncertainty of 6-10⁻⁴.



Photon emission intensities

The absolute photon emission intensities were derived from gamma- and X-ray spectrometry using HPGe calibrated detectors, with different measuring conditions to cross-reference the results.



1E8 –





The detection efficiency was calculated taking into account the asymmetry of the photomultiplier tubes, by considering the three individual TDCR values, *i.e.* T/AB, T/BC and T/AC. The absorption of the photons was calculated by Monte Carlo simulation with PENELOPE

The relative standard uncertainty due to the source dispersion was 0.13% and the global uncertainty calculation was achieved by a global Monte Carlo simulation, following the recommendations of the supplement 1 of the Guide to the expression of uncertainty in measurement.

Reference activity: 9 472 (24) kBq.g⁻¹

(Reference date: 19/12/2017, 12h00 UTC)

Half-life

The half-life of ¹⁰³Pd was measured by gamma-ray spectrometry on a high-purity germanium (HPGe) detector, using an aliquot of the diluted solution, for 50 days. The measurement was carried out by following the ¹⁰³Pd main line (357.43 keV).

Authors	Year	Half-life (d)	Absolute uncertainty (d)	Relative uncertainty (%)	Number of half-lives
Rietjens et al.	1954	17.5	0.5	2.86	
Panontin et a.	1968	16.9	0.1	0.59	
Grunditz et al.	1969	18.4	0.5	2.72	4
Czock et al.	1975	16.961	0.016	0.09	2
Vaninbroukx et al.	1981	16.991	0.019	0.11	2-3
Popov et al.	2004	16.8	0.6	3.57	5.9
Present study	2017	17.106	0.058	0.34	2.9

Spectrum of ¹⁰³Pd in the high energy range (8 keV - 550 keV)

Fit of two Gaussian functions to the experimental data of the 62-63 keV doublet

For each energy E_i , the photon emission intensity, I_i , is derived from the count rate in the relevant peak, n_i :

 ε_i : FEP efficiency for energy E_i ,

A: source activity (Bq),

 $C_{i,j}$ represents different correction factors.

	LARA (20	019)	Present study		
Energy (keV)	Photon emission intensity (%)	Relative uncertainty (%)	Photon emission intensity (%)	Relative uncertainty (%)	
XL	8.73 (23)	2.6	8.61 (43)	5	
20.07 (Κα 2)	22.05 (22)	1.0	19.59 (21)	1.1	
20.22 (Κα 1)	41.70 (40)	1.0	39.48 (42)	1.1	
K alpha total	63.75 (46)	0.7	59.07 (47)	0.8	
22.81 (Κβ1)	11.34 (15)	1.3	10.27 (11)	1.1	
23.20 (Κβ2)	1.88 (7)	3.7	1.781 (19)	1.1	
K beta total	13.22 (17)	1.3	12.05 (11)	0.9	
K total	76.97 (49)	0.6	71.1 (6)	0.9	
39.756	0.0698 (35)	5.0	0.0647 (7)	1.1	
53.285	7.2 (20) E-6	28	< 8 E-6	-	
62.41	7.8 (23) E-4	29	0.001128 (16)	1.4	
241.88	4.9 (5) E-7	10	-	-	
294.98	0.00297 (24)	8	0.00315 (7)	2.3	
317.72	1.50 (19) E-5	13	1.37 (17) E-5	13	
357.43	0.0246 (6)	2.4	0.02486 (17)	0.68	
443.8	1.60 (12) E-5	7.5	2.1 (8) E-5	33	
497.08	0.00411 (11)	2.7	0.00439 (7)	1.6	

Absolute photon emission intensities in the decay of ¹⁰³Pd

Inconsistencies in

the photon emission intensities Question on the decay scheme equilibrium

Conclusions and perspectives

The 357.43-keV photon emission intensity is 0.02464 (16) per 100 disintegrations, in agreement with the tabulated data and obtained with a lower uncertainty. In addition, the emission intensities of four other gamma rays are also derived with a significant reduction of the uncertainties.

Three gamma emissions (62.43 keV, 295.00 keV and 497.08 keV) from the third, fourth and fifth excited levels have significantly higher intensities than the tabulated data. Conversely, for the emission from the metastable level (39.75 keV), the present result is 7% lower than the tabulated value. Also, the L- and K-X-ray emission intensities are weaker (around 10%) than the tabulated data.

This would suggest that the intensity of the electron capture towards the first excited level should be weaker, while the electron capture branches to levels 3, 4 and 5 should be more intense. Another possibility would be an electron capture branch towards the stable level.

The present results provide new information and should give some useful clues to be exploited in a future evaluation of the ¹⁰³Pd decay scheme.