

Molecular spectroscopy, an enrichment diagnostic.

Laurent Bruel, Pascal Lemaire, A. Pailloux

▶ To cite this version:

Laurent Bruel, Pascal Lemaire, A. Pailloux. Molecular spectroscopy, an enrichment diagnostic.. 15th International Workshop on Separation Phenomena in Liquids and Gases (SPLG 2019), May 2019, Wuxi, China. cea-03255371

HAL Id: cea-03255371 https://cea.hal.science/cea-03255371v1

Submitted on 9 Jun 2021

HAL is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers.

L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.



{}

Proceedings of the 15th International Workshop on Separation Phenomena in Liquids and Gases

第15届"液体和气体中的分离现象" 国际会议论文集

Edited by Shi Zeng and Mingsheng Zhou 曾实 周明胜 主编

SPLG

~63~

Proceedings of the 15th International Workshop on Separation Phenomena in Liquids and Gases

第15届"液体和气体中的分离现象" 国际会议论文集

Edited by Shi Zeng and Mingsheng Zhou 曾实 周明胜 主编

"液体和气体中的分离现象"国际会议每两年召开一次,为同位素分离领域的高水平学术会议之一。本书为第15届 "液体和气体中的分离现象"国际会议的科技报告文集。本书主题针对同位素分离的基础理论和技术,重点在同位素分离 的离心法和激光法,也包括其他分离方法和技术、同位素的应用等,反映了同位素分离领域中的理论研究和应用研究的新 思想、新进展。

本书可供从事同位素分离的科研技术人员和高等院校相关专业的教师以及研究生阅读参考。

版权所有,侵权必究。举报: 010-62782989, beiqinquan@tup.tsinghua.edu.cn。

图书在版编目(CIP)数据

第 15 届"液体和气体中的分离现象"国际会议论文集=Proceedings of the 15th International Workshop on Separation Phenomena in Liquids and Gases: 英文/曾实,周明胜主编. 一北京:清华大学出版社,2020.11

ISBN 978-7-302-56813-1

I. ①第··· Ⅱ. ①曾··· ②周··· Ⅲ. ①气液分离-国际学术会议-文集-英文 Ⅳ. ①TF11-53

中国版本图书馆 CIP 数据核字(2020)第 219426 号

责任编辑: 袁 琦 封面设计: 何凤霞 责任校对: 刘玉霞 责任印制:

出版发行:清华大学出版社

址: http://www.tup.com.cn, http://www.wqbook.com

址:北京清华大学学研大厦 A 座

社 总 机: 010-62770175

投稿与读者服务: 010-62776969, c-service@tup.tsinghua.edu.cn

质量反馈: 010-62772015, zhiliang@tup.tsinghua.edu.cn

印刷者:

装订者:

销:全国新华书店 经

开 本: 210mm×285mm 印 张:18

次: 2020年11月第1版 版

定 价: 298.00 元 编:100084

HARA

购: 010-62786544

数:508 千字

次: 2020年11月第1次印刷

产品编号: 087093-01

Proceedings

The 15th International Workshop on Separation Phenomena in Liquids and Gases May 13-17, 2019 Wuxi, China

Edited by Shi Zeng and Mingsheng Zhou Professor Department of Engineering Physics Tsinghua University Beijing 100084 China

Publisher: Tsinghua University Press Printed in China





© Copyright:

All rights are reserved by the authors. No part of the papers in the proceedings may be reproduced in any form without permission of the corresponding author(s).

Publisher's Statement:

The proceedings are just a collection of the papers presented on the 15th International Workshop on Separation Phenomena in Liquids and Gases (SPLG 2019) and are meant for academic information exchange and study. That a paper is published in the proceedings by no means implies that either the publisher or the editor is for or against the point of views of the author(s), or anything that may be inferred from the paper. All papers, for which the full responsibility is taken by the author(s), are published as they are.



FOREWORD

Another SPLG Workshop (the 15th International Workshop on Separation Phenomena in Liquids and Gases, SPLG 2019) was held in China during May 13-17, 2019, 13 years after the 9th workshop (SPLG 2006) and 25 years after the 4th workshop (SPLG 1994). SPLG 2019 was jointly organized by Department of Engineering Physics and Wuxi Research Institute of Applied Technologies, Tsinghua University.

The SPLG workshop is a biannual event and has a history of 44 years, if taking the history of its predecessor into account. The predecessor of SPLG is GSR (Workshop on Gases in Strong Rotation), which was organized six times, respectively, in Stockholm, Sweden (1975), Cadarache, France (1977), Rome, Italy (1979), Oxford, UK (1981), Charlottesville, USA (1983), and Tokyo, Japan (1985). Then, The name of GSR was changed to SPLG after the 6th GSR. So far, fifteen SPLG workshops have been successfully held in Darmstadt, Germany (1987), Versailles, France (1989), Charlottesville, USA (1992), Beijing, China (1994), Iguaçu, Brazil (1996), Nagoya, Japan (1998), Moscow, Russia (2000), Oak Ridge, USA (2003), Beijing, China (2006), Angra dos Reis, Brazil (2008), St. Petersburg, Russia (2010), Paris, France (2012), San Carlos de Bariloche, Argentina (2015), Stresa, Italy (2017), and now Wuxi, China (2019). The traditional topics of the workshop had been isotope separation, but were gradually extended to a wider range of fields to include other subjects of researches concerning various separation phenomena, separation methods and techniques, providing a platform of exchanging scientific ideas and information for the people of the separation society. A very good review of the SPLG history was given by Professor Wood in his talk "History of the SPLG Workshops" in SPLG 2017. The previous GSRs and SPLGs are summarized in the following table:

Year	Sites	Participants	Host	Countries	Papers	Sessions
1975	Stockholm	37	20	9	22	5
1977	Cadarache	44	16	9	20	4
1979	Rome	70	22	10	28	6
1981	Oxford	70	15	10	26	6
1983	Charlottesville	67	40	10	24	6
1985	Tokyo	90	66	8	22	5
1987	Darmstadt	65	24	13	26	7
1989	Versailles	72	27	15	32	6
1992	Charlottesville	38	13	8	22	6
1994	Beijing	39	20	8	22	6
1996	Iguaçu	38	18	7	30	6
1998	Nagoya	90	62	9	48	6 + 3(para)
2000	Moscow	95	61	13	44	6 + 1(post)
2003	Oak Ridge	41	22	9	21	6
2006	Beijing	52	20	8	29(Oral)17(post)	8 + 1(post)
2008	Angra dos Reis	60	34	11	26	7
2010	St. Petersburg	102	66	16	53	5 + 1(post)
2012	Paris	107	40	16	37	9
2015	Bariloche	145	92	12	36	8 + 2(post)
2017	Stresa	63	18	13	35	7

The current workshop intended to cover seven topics:

- 1. Isotope Separation by Centrifugation;
- 2. Laser Isotope Separation;
- 3. Other Methods and Technologies of Isotope Separation;
- 4. New Methods and Applications of Isotopes and Their Acquisition;
- 5. Prospects and Economic Issues of Separation Methods and Technologies;
- 6. Nuclear Nonproliferation and Safeguards;
- 7. Purification Methods and Technologies.

It attracted 103 participants from nine countries. We were glad to see that there were many young researchers presenting their results, in addition to the older experienced participants with outstanding achievements. There were 30 oral presentations and 14 poster presentations, including seven plenary (special) talks given by, respectively, J. S. Tian, Z. H. Qiu, R. Snyder, V. D. Borisevich, T. Mashimo, A. Lamagna, and J. Jakutis Neto. Some excellent presentations given at the workshop are unable to be collected in the proceedings, unfortunately, because the authors were very strict about their results. We hope that they will bring their satisfactory results in the next SPLG workshop.

On behalf of the International Organizing Committee and the Local Organizing Committee, we would like to thank the strong support from Atomic Energy Industry Corporation, CNNC. We would also like to thank all participant for attending this workshop during their busy time. We are very grateful to Prof. N. N. Chen, academician of Chinese Academy of Engineering, Mr. J. S. Tian, deputy chief engineer and technique officer of CNNC, and Mr. Z. H. Qiu, chief technical officer of Atomic Energy Industry Corporation of CNNC, who spared their busy time to attend the workshop.

Prof. Shi Zeng Chairman Department of Engineering Physics Tsinghua University Beijing

Prof. Mingsheng Zhou

September 30, 2019

China

INTERNATIONAL ORGANIZING COMMITTEE

湖红

- Prof. Sergey Bogovalov, National Research Nuclear University (MEPhI), Moscow, Russia
- Dr. Diane Fischer, IAEA, Vienna, Austria
- Dr. Alberto Lamagna, Argentina National Commission of Atomic Energy, Buenos Aires, Argentina
- Dr. Pascal Lemaire, CEA, Bagnols sur Ceze, France
- Dr. Giovanni Mercurio, Joint Research Center, European Commission, Ispra, Italy

- Dr. Sylvana Presti Migliavacca, IPEN/CNEN-SP and CTMSP, Sao Paulo, Brazil
- Dr. Agnes Pailloux, CEA/DIF/DASE, Arpajon, France
- Dr. Michael Whitaker, Oak Ridge National Laboratory, Oak Ridge, USA
- Prof. Houston Wood, University of Virginia, Charlottesville, USA
- Prof. Shi Zeng, Tsinghua University, Beijing, China

LOCAL ORGANIZING COMMITTEE

- Prof. Dongjun Jiang, Department of Engineering Physics, Tsinghua University, Beijing, China
- Prof. Yongbin Jiang, Wuxi Research Institute of Applied Technologies, Tsinghua University, Wuxi, China
- Ms. Yunyan Wang, Wuxi Research Institute of Applied Technologies, Tsinghua University, Wuxi, China
- Prof. Shi Zeng, Department of Engineering Physics, Tsinghua University, Beijing, China
- Ms. Rui Zhao, Wuxi Research Institute of Applied Technologies, Tsinghua University, Wuxi, China
- Prof. Mingsheng Zhou, Department of Engineering Physics, Tsinghua University, Beijing, China

CONFERENCE CHAIR

Prof. Shi Zeng, Department of Engineering Physics, Tsinghua University, Beijing, China Prof. Mingsheng Zhou, Department of Engineering Physics, Tsinghua University, Beijing, China

ORGANIZERS

Department of Engineering Physics, Tsinghua University Wuxi Research Institute of Applied Technologies, Tsinghua University

SUPPORTERS

Joint Research Institute for Nuclear Fuel Cycle and Materials,

Tsinghua University—China Nuclear Energy Industry Corporation
China Nuclear Energy Industry Corporation, CNNC

CONTENTS

$I \ . \ Laser \ Technology$

Molecular Spectroscopy, an Enrichment Diagnostic (L. Bruel, P. Lemaire, A. Pailloux)
Advances in the Argentine Project of Uranium Isotope Separation by Laser Technology
(A. Lamagna, P. D. Fainstein, D. Fregenal, P. Knoblauch, M. Zarco, J. D. Fuhr, J. Fiol,
E. Cortizo, E. Kaul, E. Santos, J. Gonzalez, N. Pachao, G. Berlín, S. Clavijo, E. Kirchuk)
Selectivity in Studies of Laser-induced Isotope Separation in Pulsed Free Expansions of SF ₆ and
UF ₆ (D. Fregenal, P. Knoblauch, M. Zarco, E. Cortizo, P. D. Fainstein, J. D. Fuhr,
E. Kaul, A. Lamagna, P. Maceira, E. Zemma, J. Fiol)
Spatial Distribution of Dysprosium Vapor Generated by Laser Ablation as a Function of
Pulse Energy (J. Jakutis Neto, P. Bueno, A. A. Almeida, C. A. B. Silveira, K. K. Yum,
J. W. Neri)
Modelling for Estimating Total Ion Extraction Time of a Barium Plasma Under a Parallel-plate
Collector Configuration (J. Chen, J. Li, H. P. Li, D. J. Jiang, M. S. Zhou)
Two-dimensional Simulation of Near-longitudinally Pumped Dye Laser Amplifiers
(Q. N. Yang, X. W. Zhang, Z. J. An, S. Zeng) 40
Research on Light-induced Drift Alkali Metal Isotope Separation (J. Q. Yang, T. Liu,
J. D. Liu, C. Jin, Z. Z. Zhang)
The Rubidium Evaporation Research in Rubidium Isotope Separation Through the
Photochemical Method (T.G. Yang, J.G. Jiang, S. W. Deng, L. H. Zhou)
Dependence of the Separative Power of an Optimized Iguassu Centrifuge on the Rotor Parameters
(S. V. Bogovalov, V. D. Borman, V. D. Borisevich, I. V. Tronin, V. N. Tronin)
Study of the Circulation Flow in a Centrifuge During the Enrichment of Isotope
Mixture of Heavy Gases (E. P. Potanin, L. Y. Sosnin, A. N. Cheltsov)
A Method for Analyzing the Transient Hydraulics in a Gas Centrifuge (S. Zeng, Z. G. Lei,
V. D. Borisevich, A. Y. Smirnov, G. A. Sulaberidze)
Numerical Study on the Effect of the Feed Jet in Gas Centrifuge (Z. Y. Gu, D. J. Jiang)
Study on Effects of Rotor Wall Pressure and Withdrawal Flow in a Gas Centrifuge
(H. R. Chen, D. J. Jiang, S. Zeng)
Study on the Disc Withdrawal Chamber with the Molecular Pump Inside the Rotor for
Gas Purification (G. Pei, Y. B. Wang, D. J. Jiang, M. S. Zhou)
Study on Separative Performance of Single Centrifuge with N-Octane as Medium
(Q. M. Sun, J. X. Pan, C. H. Jian, G. Pei, M. S. Zhou)

Numerical Study of the Hydraulic Characteristics of the Bi-directionally Connected Gas	
Centrifuge (M. An, Y. N. Zhang, S. Zeng, Z. G. Lei, V. D. Borisevich, A. Y. Smirnov)	110
Study on Performance of Internal and External Holweck Pumps (N. Y. Ling, D.J. Jiang,	
Y. Sheng, X. Lu, C. L. Zhou, X. H. Li)	117
Study of Numerical Simulation Method of the Radial Separation Effect in the High-speed	
Rotating Flow Field (Y. Sheng, X. Lu, C. L. Zhou, N. Y. Ling)	121
Research on On-line Active Control Mechanism of Rotor Sidewall Temperature	
(N. Y. Ling, X. K. Chen, L. Wang, R. Sun)	128
3D Numerical Study on the Gas Flow Field and Concentration Field in a Strong	
Rotating Cylinder (X. H. Li)	132
A New Type of Plasma Centrifuge for Isotope Separation (V. D. Borisevich,	1 41
E. P. Potanin, J. V. Whichello)	141
Electro-induced Separation of Solvated Ions in Solutions (I. V. Shamanin, M. A. Kazaryan,	1.47
S. N. Timchenko, I. A. Ushakov, A. D. Poberezhnikov, H. D. Li)	147
Study on Boron Isotope Separation by Gas Diffusion Method Using BF ₃ as Processing Gas (Y.B. Wang, G.Pei, M.S.Zhou)	156
(1.B. Wang, G. Fei, M.S. Zhou)	130
Separation of Multicomponent Isotope Mixture During Filling of Gas Centrifuge	
Cascade (A. A. Orlov, A. A. Ushakov, V. P. Sovach)	165
Reprocessed Uranium Enrichment for Sustainable Nuclear Fuel Supply	
(A. Y. Smirnov, V. E. Gusev, G. A. Sulaberidze)	175
Obtaining Highly Enriched Intermediate Isotopes in Cascades of Gas Centrifuges with	
Additional Product Flows (A. Y. Smirnov, D. N. Fomin, S. Zeng, G. A. Sulaberidze,	
A. R. Mustafin, V. D. Borisevich, D.J. Jiang)	183
Economic Analysis of High Abundance ¹⁰ B Isotope Production by Centrifugal Method	100
(C. H. Jian, J. X. Pan, Q. M. Sun, G. Pei, M. S. Zhou)	190
Technology Research for Separation of Silicon-28 Isotope Production by a Gas Centrifuge	
(D. D. Dou, Y. K. Cong)	196
Study on the Effect of Tail Abundance on Centrifugal Cascade Efficiency and	100
Its Control Method (Z. Y. Zhang, H. W. Wu)	203
Application of Genetic Algorithm in Cascade Optimization (H. T. Yao, M. Zhang, Y. Liao)	
Research on the Calculation Method for Improving Conformity of the Cascade Scheme	200
(F. Zou, J. Chen, T. Pan, L. Z. Zeng)	219
Definition of Static Characteristic Roots of a Centrifugal Cascade (W.J.Li,	213
Z.D. Wu, H. Yan)	225
Application of Simulation in Cascade Fault Diagnosis (H. Yan, W.J. Li, C.Q. Ge)	
Study Measurement and Optimization of Operation Parameters of a Centrifugal	۷۵۷
Cascade (J. X. Gan, B. Oiu)	220
Vascauc (J. A. Ually D. Olu)	439

CONTENTS • ₩ •

Analysis of Response and Factor Correlation Based on Complex Equipment	
Performance Data (C.L.Zhou, X.Lu, Y.Sheng, R.L.Wang)	248
Study on Unsteady State of Low Abundance Centrifugal Cascade (Z. H. Liu, Z. Y. Zhang)	258
Numerical Study on the Process of UF ₆ Desublimation in a Tank (Y.N.Feng, D.J.Jiang)	264
List of Participants	270
Author Index	275



I. Laser Technology

Molecular Spectroscopy, an Enrichment Diagnostic

L. Bruel^{1, *}, P. Lemaire¹, A. Pailloux²

- Department of Enrichment Technologies, Dismantling & Wastes
 Atomic Energy Commission, France
 * Corresponding author, Email: laurent.bruel@cea.fr
 - 2. Department Analysis Surveillance, Environment Atomic Energy Commission, France

Abstract

In Molecular Laser Isotope Separation processes, high-resolution spectroscopy is commonly used to determine the rotational and vibrational temperatures.

In both versions of the MLIS processes (dissociation or SILEX version), the jet-cooled gas tends to form rapidly van der Waals complexes with the rare gas used as carrier gas. These phenomena are important for isotopic selection as they initiate the condensation phenomena.

In case of SF_6 , small heteroclusters (SF_6 -rare gas) are investigated with rovibrational laser jet-cooled spectroscopy in the ν_3 region (10 μ m for SF_6), for some rare gas atom (RG). High-resolution spectra are measured for both the perpendicular and parallel bands for Ne, Ar, Kr and Xe heteroclusters. It is shown that the reduced vibrational red shifts are independent of the RG atom, leading us to a good understanding of their conformation.

In the case of UF $_6$, the unusual wavelength of the ν_3 band required the development of a specific DFB laser at 16 μm . This has been done at IES (CNRS/University of Montpellier) using the InAs/AlSb quantum cascade laser (QCL-DFB) technology that demonstrated room temperature operation in continuous regime for the first time above 16 μm with a linewidth of a few MHz.

Absorbance measurements with these lasers are now demonstrated on isolated NH $_3$ lines in the 16 μm range, at low pressure. Thermal induced frequency modulation (2f/f) is used to improve signal-to-noise ratio.

In the case of online isotopic ratios measurements, this laser could measure the UF₆ molecule number density of both isotopologues. For that purpose, it will be coupled to an integrated pulsed supersonic jet. **Keywords:** SPLG 2019, UF₆, SF₆, Complex, Dimer, HTDS, Isotopy, Quantum Cascade Laser

1. Introduction

Several teams have studied the absorption spectrum of the ν_3 vibration of UF₆. Takami et al. have even published very detailed spectral portions to isolate the Q branch of $^{235}UF_6$ between the R(5) and R(6) lines of ²³⁸UF₆ in a pulsed supersonic jet. To avoid hot bands (multi excited vibration with v_3) and bring all the molecules into the ground state, the UF₆ plus buffer gas mixture should be cooled to below 50 K in an adiabatic expansion. At these tempera-tures van der Waals complexes appear. They are the product of homogeneous or heterogeneous aggregation of UF₆, with the buffer gas, marking the beginning of condensation. The spectroscopy technic that we have implemented allows us to identify the different states of the gas in the expansion, either on a simulant of the molecular structure, the SF₆, or on the UF₆, according to the available Quantum Cascade Laser (QCL) diodes (10 μ m or 16 μ m). We report here the detailed spectro-scopic results for SF₆, conducted with a pulsed gas expansion, and also a continuous jet. The first paragraph focuses on aerodynamicresults—temperature measurements—in continuous relaxa-tion. The next identifies dimers and clusters bands versus the buffer gas type. The third will deal with QCL performance at 16 μ m. The last paragraph will conclude the use of this diagnostic in isotopic separation.

High-resolution spectroscopy of UF₆ for aerodynamic

2.1 Theoretical calculations

To perform a sufficient spectral identification in the infrared region within a broad spectral range at very high speeds, only the absorption spectroscopy with QCL laser diode is possible. Now that it is feasible, it was necessary to be able to compare the spectra with the theory to deduce very accurately the vibrational and rotational temperatures. Fortunately, the cold band of UF_6 can be calculated using Highly-spherical Top Data System (HTDS) software^[2].

Calculated temperature dependence of the spectra is very strong for UF₆ as shown on Figure 1 and on Figure 2, for respectively 20 K and 10 K; on Figure 3 is shown the superposition of 235 UF₆ and 238 UF₆ at 1 K. Even with an isotopic shift of 0.65 cm $^{-1}$, the two sets of rotational lines are not separated. Thanks to highresolution spectroscopy, most of the rotational lines can be separated.




Figure 1: 235 UF $_6$ and 238 UF $_6$ cold bands at 20 K




Figure 2: 235 UF $_6$ and 238 UF $_6$ cold bands at 10 K




Figure 3: In orange, 235 UF, rotational lines are inside the rotational structure of 238 UF $_6$, in blue

2.2 Rotational temperature measurements

When we use SF₆ to simulate UF₆ we must be aware that the isotopic shifts are really different, as you can see on Figure 4 where all sulfur hexafluoride isotopologues are separated at all temperatures. Nevertheless, SF₆ is a validating tool to interpret phenomena relative to either aerodynamic or aggregation.




Figure 4: ν_3 vibration of the 34 SF $_6$ 33 SF $_6$ 32 SF $_6$ isotopologues respectively

The way to get very good temperature measurements is as follows. A comparison theory/experiment is performed for all rotational transitions (Figure 5) and the best fit gives the rotational temperature. On this graph, realized with a laser resolution of 20 MHz, the calculated (HTDS) and measured spectra are in a very good agreement even at the line level. The rotational temperature can then be deduced with a precision of less than $\pm 5 \text{ K}.$



Figure 5: Theory and experiment agreement on the R side of the ν_3 vibration of SF_6 , at 20 MHz resolution

2.3 Vibrational temperature measurements Hot bands involve other vibrational levels depending

on vibrational temperature. Hot bands appear as the vibrational temperature increases. We use them to

measure that temperature, as hot bands population is linked to it. Assuming Boltzman distribution, the two Q branch surfaces can give an estimation of the expected temperature.




Figure 6: ν_3 bands at 100 K vibrational temperature and at 20 K rotational temperature, hot bands (in red) are visible (for instance here involving ν_5 bands vibration at 946, 9 cm⁻¹)

3. SF₆-Rg spectroscopy (clusterization)

The different configurations of aggregates (homoclusters, heteroclusters) in the different aerodynamic regimes, partial concentrations or gas laser interactions must be understood to choose the operating point of a laser isotopic separation process.

Aggregate high-resolution spectroscopy has been successfully investigated for the ν_3 vibration with the pulsed jet experiment of MONARIS in Jussieu

University [3]. Experiments have been carried out with different rare gases (Rg).

3.1 SF₆-Rg heteroclusters

The parallel and perpendicular bands of SF_6 -Rg heteroclusters have been identified. These bands arise from the degeneracy lifting of the triply degenerate state (ν_3) of the SF_6 monomer.

These experimental spectra are very weak compared to the strong monomer band. But, thanks to the accuracy of the calculated monomer band, the heteroclusters band is obtained by subtracting the theoretical spectrum from the measurement (in green on Figure 7).

These cluster spectra have been successfully computed and the experimental spectral shifts are in good agreement with the prediction based on a Buckinghamtype potential. Some molecular parameters of this model (rotational constants, band centers and S-Rg distances) have been deducted from the measurements^[3].

$3.2 (SF_6)_2$ -dimer

The first conformer of $(SF_6)_2$, (band #1), has been published in 1995 by Urban and Takami. A second conformer (band #2) has been identified recently^[3].




Figure 7: EC-QCL jet cooled pin hole spectra of both ν_3 parallel bands of \overline{SF}_6 -Rg heterodimers (Rg=Ne, Ar, Kr, Xe) with 0.12% \overline{SF}_6 and 1% Rg diluted in 8 bar helium at an axial distance z=15 mm

Again, with the experimental data, they have been simulated by symmetric top transition, taking into account their symmetries D_{2d} and C_{2h} and their temperatures of 3 K and 8 K respectively (Figure 8).

Temperatures are distinct because # 2 is formed during early expansion at higher temperature. A collisional relaxation process leads to population conversion from dimer # 2 to dimer # 1.




Figure 8: Parallel bands of the conformations #1 and #2 of the dimer $(SF_6)_2$, summation compared to experimental spectrum

3.3 (SF₆)₂-Rg heteroclusters

When the structure of the homodimer SF_6 was well understood, the continuation of the heterogeneous aggregation in the mixture supersonic expansion was studied with several rare gases, for different geometries (slit or nozzle), SF_6 concentration, axial probe distance.

The #3 band (in red Figure 9) was easily assigned to the heterocluster $(SF_6)_2$ -Rg, because of linear dependence of the vibrational shift between #2 and #3 bands with gas polarizability.

This linear dependence is showing that the trimer #3 configuration is the same as #2. This illustrates the lack of robustness of the #1 dimer, owing to external perturbations, like collisions or complexation with a He atom.

Finally, these results show that it is necessary to use sophisticated approaches to simulate these systems [4].




Figure 9: Spectra of different heteroclusters $(SF_6)_2$ -Rg for Rg=He to Xe, linear dependence of the vibrational shift with the polarizability of the rare gas atom is shown

4. QCL a tool for high-resolution spectroscopy

All these experiments have been done using the pulsed jet SPIRALES bench in Paris or on the continuous jet in our lab, with the same external cavity Quantum Cascade Laser at a wavelength of $10\mu m$ from a commercial supplier.

The energy domain of the vibrations and the isotopic differences are weaker for UF₆, which will modify the behavior of the aggregates in the relaxation, the high-resolution spectroscopy will make it possible to understand the behavioral changes. To investigate the high-resolution spectroscopy of UF₆, there is (in 2019) no commercial device available at the wavelength of 16 μ m.

Thanks to a collaboration with the Institute of Electronics and Systems (IES, CNRS/University of Montpellier) in France, the the first world CW 16μm laser source^[5,6] operating up to ambient temperature has been developed in the University of Montpellier. We used one of these sources in this study. This laser can accept not only continuous regime, but it also tolerate room temperature regime at the same time. It can deliver 50 mW CW for approximately 7 W of electrical power.

QCLs are different from laser diode (Figure 10) as they use many current injectors in series connected by a quantum tunneling. Each injector provides one photon from the same electron. The gain can be high, depending on materials and design.

The wavelength is first obtained by adjusting the design of the gain structure which is an epitaxed multilayer, with a Bragg grating etched on top on the ridge to obtain a monomode emission linewidth of a few MHz.

InAs is advantageously used at higher wavelength, thanks to its electron's low effective mass, and it has been extended in the far IR (up to 25 μ m). On the contrary InP is not suitable because the substrate is absorbing very near to 16 µm, due to a phonon resonance.

QCL Compared to Diode Lasers




Figure 10: AlSb / InAs active zone (AlSb tunnel ~1 Å to 100 Å), One period=20 MoCVD layers Np=144 periods (7 um thick)

4.1 The use of QCL at 16 μ m

Due to thermal effects, the emitted wavelength depends on both temperature and current. It is then possible to scan a 1.5 cm⁻¹ range by varying the current, and scan a 6 cm⁻¹ range by changing the operating temperature, as shown on Figure 11.




Figure 11: Example of current tunablity at fixed temperature of our QCL ship

4.2 Laser test on NH₃ gas cell

A simple teston a low-pressure ammonia cell has been conducted to confirm that this laser source is usable for UF₆ spectroscopy.

Linewidth was found less than 30 MHz and the laser could effectively scan from 627 to 634 cm⁻¹ by varying the temperature (Figure 12).

4.3 2f / f detection from laser modulation

In addition to a small size, a large power at $16 \mu m$, a very high spectral resolution..., the QCL can be modulated at a maximal frequency of 100 kHz. A synchronous demodulation with detection of the first and second orders permits us to increase signal to noise ratio, using the so-called 2f/f technic.




Figure 12: Experimental scan from 627 to 634 cm $^{-1}$ with a 16 μm QC Laser. The red arrow indicates where the central position of the $^{235}\,\mathrm{UF}_6$ lines is on the ammonia spectrum

Figure 13 is a numerical simulation with noisy are blurred, the demodulated signals are clean. signals coming from the 235 UF₆ branch HTDS calculus and modulation of the laser. Then detection of the first and second orders is undertaken.

Compared to direct absorption, where the signals

This technic is a necessary tool to obtain the high dynamic needed to acquire both isotopes at low assay.




Figure 13: Simulation of noisy signals demodulated using the 2f/f scheme coming from the 235 UF $_6$ branch (red and green) to be compared to direct absorption (in grey)

5. Conclusions

MLIS-type isotopic separation techniques are known to be potentially profitable, that is to say with high selectivity and with access to the entire UF₆ population, only involving isentropic expansion carrying the gas mixture at very low temperature. However, during the relaxation, a dimerization followed by homogeneous or heterogeneous aggregation arises according to the conditions of the process. To properly control the thermodynamic state of the gas mixture, spectroscopy proves here R. to be the best tool. At high resolution it discriminates pell the types of aggregates, and for a process monitoring, UM it discriminates the isotopes, even if the rovibrational spectra are nested. For UF₆, the wavelength of the characteristic vibration is in the far infrared and the only tool with this discriminant potential is the QCL [1] laser diode that we used in this study.

Acknowledgements

This work has only been possible thanks to the following people and institutes:

M. A. Gaveau, M. Mons, V. Brenner(LIDYL, CEA, CNRS, *Université Paris-Saclay*, CEA Saclay, 91191 Gif-sur-Yvette, France)

P. Asselin (Sorbonne Universités, UPMC Univ. Paris 06, UMR 8233, MONARIS, F-75005, Paris, France)
V. Boudon, (Laboratoire Interdisciplinaire Carnot de Bourgogne, UMR 6303 CNRS Université)

A. C. Turner (Department of Earth and Planetary Science, University of California, Berkeley, California 94720 USA)

A. Potapov (Laboratory Astrophysics Group of the Max Planck Institute for Astronomy at the Friedrich Schiller University Jena, Germany)

R. Teissier, A. N. Baranov (CNRS & Univ. Montpellier, IES (Institute of Electronics and Systems), UMR 5214, F-34000, Montpellier, France)

References

- [1] Urban R D, Takami M J. Chem. Phys., 1995, 102, 3017.
- [2] Wenger C, Boudon V, Champion J P, et al. Highly-spherical Top Data System (HTDS) software for spectrum simulation of octahedral XY6 molecules[J]. Journal of Quantitative Spectroscopy and Radiative Transfer, 2000, 66(1): 1-16.
- [3] Asselin P, Turner A C, Bruel L, et al. Rovibrational laser jet-cooled spectroscopy of SF_6 -rare gas complexes in the ν_3 region of $SF_6[J]$. Physical Chemistry Chemical Physics, 2018, 20(44).
- [4] Asselin P, Potapov A, Turner A C, et al. Conformational landscape of the SF6 dimer as revealed by high resolution infrared spectroscopy and complexation with rare gas atoms[J]. Phys., Chem. Chem. Phys., 2017; 17224-17232.
- [5] Bahriz M, Lollia G, Baranov A N, et al. High temperature operation of far infrared ($\lambda \approx 20 \mu m$) InAs/AlSb quantum cascade lasers with dielectric waveguide[J]. Optics Express, 2015, 23(2).
- [6] Baranov A N, Bahriz M, Teissier R. Room temperature continuous wave operation of InAs-based quantum cascade lasers at 15μm[J]. Optics Express, 2016, 24(16).