DEM 2021 – International Conference on Decommissioning Challenges: Industrial Reality, Lessons learned and Prospects

France, Avignon - 2021, September 13 | 15

Determination of CI-36 in various radioactive waste

Céline Gautier^{1*}, Christèle Colin¹, Margaux Giuliani¹, Elodie Laporte¹, René Brennetot¹, Laëtitia Kasprzak¹, Pascal Fichet¹

 Université Paris-Saclay, CEA, Service d'Études Analytiques et de Réactivité des Surfaces, 91191, Gif-sur-Yvette, France.
* Corresponding Author, E-mail: celine.gautier@cea.fr

KEYWORDS: radioactive waste, radiochemical method, CI-36

Introduction

In the context of radioactive waste management, Cl-36 is a key radionuclide to be measured due to its long half-life (3.01x10⁵ years) and its high mobility in the environment. In particular, its mass activity must be lower than 5 Bq/g in France in the packages stored in the disposal facility dedicated to low and intermediate level nuclear waste. Presence of Cl-36 in radwaste mostly results from neutron activation of stable chlorine which can be contained as an impurity in fuels, fuel claddings, graphites and construction materials (steels and concretes).

Different strategies developed to extract and purify CI-36 in radioactive waste

CI-36 decays by emitting a beta particle (98.1 % yield - E_{max}=708.6 keV) and electron capture (1.9 %). Consequently, this radionuclide can be measured by beta counting techniques, mainly liquid scintillation counting (LSC) for routine analysis of radwaste. As LSC has a poor energy resolution, CI-36 has to be extracted and isolated from matrix and interfering elements through a selective radiochemical method before any LSC measurement. Radiochemical procedures are generally based on chlorine extraction followed by its precipitation with silver as AgCl and dissolution of this AgCl precipitate with ammonia or implementation of ion-exchange chromatography or extraction chromatography [1]. The method based on AgCl precipitation is relevant for radwaste with low beta- and gamma-emitters and with no other halogens such as iodine (which also precipitates with silver). LASE laboratory (Operator Support Analysis Laboratory) has developed an alternative radiochemical protocol in order to achieve higher decontamination factors required for Cl-36 analysis of high level radioactive samples [2]. After chlorine extraction with various decomposition methods, the following selective purification steps are applied: chlorine precipitation with silver, organic synthesis (chlorine fixation on octadecene), purification of the chlorinated molecule, then its destruction and an ultimate liquid-liquid extraction. Cl-36 is finally measured by LSC whereas stable chlorine is analyzed by ion chromatography to determine the separation yield.

The presentation will give an overview of the different strategies developed to extract and purify CI-36 in various radioactive waste. The validation of the radiochemical method will first be performed using inhouse reference samples spiked with CI-36 and the participation in interlaboratory comparisons. Different methods applied to extract CI-36 from various radioactive matrices will then be presented. In particular, implementation of micro-wave digestion, pyrolysis or acid digestion or leaching assisted with digital autoradiography will be explained depending on the studied matrices: ion exchange resins, graphites, concretes, irradiated or surface-contaminated steels and magnesium claddings. Finally, the presentation will discuss with the adaptation of the radiochemical procedure to measure CI-36 by accelerator mass spectrometry (AMS) in order to achieve a detection limit lower than 5 Bq/g.

Conclusion

CI-36 can be determined with accuracy and high sensitivity in different types of radioactive waste. The performances achieved for CI-36 are largely suitable for the monitoring of nuclear waste in disposal facilities.

References

- [1] AFNOR. French Standard NF M60-332 Nuclear Fuel Technology Waste Determination of Chlorine-36 activity in effluents and waste by liquid scintillation counting (2010).
- [2] Frechou, C., Degros, J.P. Measurement of ³⁶Cl in nuclear wastes and effluents: Validation of a radiochemical protocol with an in-house reference sample. *J. Radioanal. Nucl. Chem.* 263 333-339 (2005).