



HAL
open science

Determination of fractional electron capture probabilities of Nickel-59 by means of Metallic Magnetic Calorimeters

Arshjot Kaur, Lucille Chambon, Martin Loidl, V. Lourenço, Matias Rodrigues

► **To cite this version:**

Arshjot Kaur, Lucille Chambon, Martin Loidl, V. Lourenço, Matias Rodrigues. Determination of fractional electron capture probabilities of Nickel-59 by means of Metallic Magnetic Calorimeters. 20th International Conference on Low Temperature Detectors, Jul 2023, Daejeon, South Korea. 2023. cea-04137584

HAL Id: cea-04137584

<https://cea.hal.science/cea-04137584>

Submitted on 24 Jul 2023

HAL is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers.

L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.


The measurement is part of the European metrology research project MetroMMC (2018-21). In the project, MMCs have been optimized and utilized for decay energy spectrometry of radionuclides decaying via electron capture (EC). Several EC-decaying radionuclides emitting radiation in the energy range between 20 eV and 100 keV were selected.

The project's main objectives are to measure fractional electron capture probabilities and absolute X-ray emission intensities. To improve theoretical models and calculations of electron capture and subsequent atomic relaxation and to validate the theoretical data with the high-precision experimental data from this project.

Motivation



Nickel-59, one of the radionuclides chosen for the MetroMMC project, decays mainly by EC to the ground state of Co-59 and has a long half-life of 76,000 years. The decay is a 2nd forbidden non-unique transition. The main applications where the electron capture probability ratios are required are:

- Isotope geology: ⁵⁹Ni is used to date meteorites and determine the abundance of extraterrestrial dust in ice and sediment.
- Nuclear power plant radioactive waste qualification: ⁵⁹Ni is produced in nuclear reactors as an activation product of Ni-cladded control rods and other steel structural components. It is present in nuclear waste, and its detection is very difficult because of its long half-life and no γ -ray emission.




Detection technique: Metallic Magnetic Calorimeters (MMCs)

- MMCs with the radionuclide embedded in the absorber \rightarrow Decay energy spectrometry (DES).




Absorber and source preparation

I. Electrodeposition Method




II. Micro-drop deposition on gold nanofoam






- Figure 4a: Au_{0.3}Ag_{0.7} (wt%) alloy (2 \times 2 \times 0.006 mm³) diffusion welded on gold foil (2.5 \times 2.5 \times 0.01 mm³).
- Figure 4b: Gold nanofoam by dealloying through wet etching with 20% HNO₃ for 2 hours at 50 °C, dissolving silver, and forming nanoporous gold.
- Figure 4c: Micro-drop dispenser setup. For a 3 Bq activity, the drop volume was 41 nL (1216 drops); for 4 Bq, it was 55 nL (1671 drops).
- Figure 4d: Autoradiography of one of the sources.

Experiment setup: Decay energy spectrometry (DES)

- In DES measurement mode, the source is fully enclosed between the two absorber halves (Figure 5a) to have a 4 π detection geometry (Figure 5b).
- The decay energy from the source, except the neutrino, is thermalized inside the absorber. The obtained energy spectrum shows a single peak for each capture electron shell.



Spectrum analysis and results



Fractional electron capture probability ratios

Measurement	P_L/P_K
Electrodeposition	0.1154 (11)
Drop deposition	0.1122 (22)
BetaShape *[3]	0.1159 (26)
Literature	0.1210 (20) [4]

* Theoretical calculation code developed at LNHB.

- The ratio of peak areas corresponds to the fractional capture probabilities. The quoted uncertainties comprise the statistical uncertainties and an estimation of the uncertainty due to the data analysis.
- The pulses had a 10-90% rise time of 170 μ s in both sources, with exponential decay resulting in a 1/e decay time of 400 μ s in the electroplated source and 440 μ s in the drop-deposited source.
- In drop-deposited sources, due to the loss of a portion of the energy of Auger electrons in the source material causes a low energy tail (0.95% for K peak), less pronounced in electroplated sources (0.38% for K peak).

The fractional electron capture probability ratios from both sources are compatible with each other and the recent BetaShape calculation value.

The follow-up EMPIR project: PrimA-LTD (2021-2024)

- To develop a new primary method for decay scheme independent activity determination, high energy resolution, and statistics that exceed 10⁸ events/spectrum.
- High statistics measurements of ¹²⁹I (beta spectrum shape; branching ratio) and ⁵⁵Fe (shake-up and shake-off effects) will be performed by the use of multiple detectors.

➤ Upgraded MMC spectrometer setup to acquire measurement statistics of >10⁸ events per spectrum.

