



**HAL**  
open science

## scCVD diamond membrane based microdosimeter for hadron therapy

Izabella Zahradnik, Philippe Barberet, Philippe Bergonzo, Ludovic de Marzi,  
Wataru Kada, Stukan Natko, Michal Pomorski, Samuel Saada, Dominique  
Tromson

► **To cite this version:**

Izabella Zahradnik, Philippe Barberet, Philippe Bergonzo, Ludovic de Marzi, Wataru Kada, et al..  
scCVD diamond membrane based microdosimeter for hadron therapy. Hasselt Diamond Workshop  
2018 - SBDD XXIII, Mar 2018, Hasselt, Belgium. pp.184. cea-04174606

**HAL Id: cea-04174606**

**<https://cea.hal.science/cea-04174606>**

Submitted on 1 Aug 2023

**HAL** is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers.

L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.

## scCVD Diamond Membrane based Microdosimeter for Hadron Therapy

Izabella Zahradnik<sup>1\*</sup>, Philippe Barberet<sup>5,6</sup>, Philippe Bergonzo<sup>3</sup>, Ludovic De Marzi<sup>4</sup>, Wataru Kada<sup>8</sup>,  
Skukan Natko<sup>7</sup>, Michal Pomorski<sup>1</sup>, Samuel Saada<sup>1</sup>, Dominique Tromson<sup>2</sup>

<sup>1</sup>CEA-LIST, Diamond Sensors Laboratory, 91191, Gif-sur-Yvette, France

<sup>2</sup>CEA-LIST, Sensors and Electronic Architectures Laboratory, 91191, Gif-sur-Yvette, France

<sup>3</sup>CEA-LIST, 91191, Gif-sur-Yvette, France

<sup>4</sup>Institut Curie, Centre de Protonthérapie d'Orsay, 91400, Orsay, France

<sup>5</sup>Université de Bordeaux, CENBG, Chemin du Solarium, 33175 Gradignan, France

<sup>6</sup>CNRS, UMR5797, CENBG, Chemin du Solarium, 33175 Gradignan, France

<sup>7</sup>Division of Experimental Physics, Ruđer Bošković Institute, 10000 Zagreb, Croatia

<sup>8</sup>Faculty of Science and Technology, Gunma University, Kiryu, Gunma 376-8515, Japan

[izabella.zahradnik@cea.fr](mailto:izabella.zahradnik@cea.fr)

Hadron therapy (protons and carbon ions) is an innovative mode of radiation therapy for cancer treatment which enables tumour cells to be more effectively destroyed than conventional radiation therapy using photons. Knowledge of the high linear energy transfer<sup>1</sup> (LET) of particles is used as a fundamental parameter in the prediction of the relative biological efficiency<sup>1</sup> (RBE) of particle beams in hadron therapy. Therefore, it is a key quantity to be considered in ensuring that a safe dose is delivered to patients in treatment planning.

Based on single-crystal CVD (scCVD) super-thin diamond membranes obtained using deep Ar/O plasma etching<sup>2</sup>, a prototype of a solid-state microdosimeter for LET measurements was produced at the Diamond Sensors Laboratory of CEA-LIST. A p<sup>+</sup>-i junction was created by a CVD growth of thin 400 nm boron-doped scCVD layer on top of an intrinsic 4 μm thick diamond membrane. Multiple square microdosimetric sensitive volumes (μSVs) with side lengths of 120, 60 and 30 μm (Figure 1) were formed on the membrane surface using projection photolithography, followed by local plasma etching of the p<sup>+</sup> diamond layer<sup>2</sup>. The characteristic behaviour of the micro p<sup>+</sup>-i junction was determined by a 1.8 V built-in potential which resulted in full depletion of the device at zero-bias.




Figure 1: Scanning electron microscope (SEM) image of the scCVD diamond membrane based microdosimeter prototype with different micro-sensitive volumes.

The response of the diamond membrane microdosimeter to single projectiles was investigated in ion micro-beams. The active μSVs of the microdosimeter were irradiated using a raster scanning method and the charge transport properties of the device were determined with sub-micron precision by measuring the charge collection efficiency (CCE), the μSVs 3D spatial definition and the pulse-height spectra. Preliminary, this novel microdosimeter was tested in a 100 MeV therapeutic proton beam at Institut Curie – Proton Therapy Center in Orsay and the experimental data compared to Monte Carlo simulation. The results show a great potential of this device to be used for studies of the RBE in clinical applications.

### References

1. E. J. Hall. *Radiobiology for the radiologist*. Lippincott Williams & Wilkins, Philadelphia, 2000.
2. M. Pomorski, B. Caylar, P. Bergonzo, *Appl. Phys. Lett.* 103, 112106, 2013.