

Self-attenuation in the low-energy range: an experimental study on 210Pb

Marie-Christine Lépy, Lucille Chambon, Benoit Sabot, M. Anagnostakis, Anne de Vismes Ott, R. Galea, R. Idoeta, P. Jodlowski, K. Karfopoulos, Anne Meyer, et al.

▶ To cite this version:

Marie-Christine Lépy, Lucille Chambon, Benoit Sabot, M. Anagnostakis, Anne de Vismes Ott, et al.. Self-attenuation in the low-energy range: an experimental study on 210Pb. ICRM 2023 - 23rd International Conference on Radionuclide Metrology and its Applications, Mar 2023, Bucarest, Romania., 2023. cea-04204552

HAL Id: cea-04204552 https://cea.hal.science/cea-04204552

Submitted on 12 Sep 2023

HAL is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers.

L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.



Title: Self-attenuation in the low-energy range: an experimental study on ²¹⁰Pb

Authors names: M.-C. Lépy¹, L. Chambon¹, B. Sabot¹, M. Anagnostakis², A. De Vismes³, R. Galea⁴, R. Idoeta⁵, P. Jodlowski⁶, K. Karfopoulos⁵, A. Meyer³, I. Mitsios², V. Peyres⁶, P. Saganowski⁶, N. Salpadimos⁵, M.I. Savva¹o, O. Sima¹¹, T.T. Thanh¹², R. Townson⁴, Z. Tymiński⁶, T. Vasilopoulou¹o, L. Verheyen¹³, T. Vidmar¹³

Affiliations (short names), 1: Université Paris-Saclay, CEA, List, Laboratoire National Henri Becquerel (LNE-LNHB), F91120, Palaiseau, France, 2: NTUA, 3: IRSN, 4: NRC, 5: UPV/EHU, 6: AGH UST, 7: EEEA, 8: CIEMAT, 9: POLATOM, 10: NCSR "Demokritos", 11:U. Bucharest, 12: VNUHCM, 13: SCK CEN

Text:

Quantitative analysis of environmental samples generally involves volume source with low radioactivity and an unknown composition of the sample matrix. Unless an efficiency calibration with the same geometry and matrix is available, it is necessary to correct the raw counting rates for self-attenuation effects in order to obtain the correct sample activity. This can be achieved either experimentally or by means of calculation based on mass attenuation coefficients that can be directly measured or taken from tabulated absorption data. The self-attenuation effect is more pronounced when the energy of the emitted photons is low. This is the case with ²¹⁰Pb, which is a low-energy gamma emitting radionuclide (46.54 keV) suffering strong absorption in the sample matrix and whose concentration in the environment must be regularly monitored according to public health regulations.

Within the framework of the GSWG, twelve laboratories agreed to carry out measurements with matrices containing ²¹⁰Pb and to compare their approaches to the determination of self-attenuation correction factors in order to draft practical recommendations for users.

With this aim, the Laboratoire National Henri Becquerel (LNHB) prepared two different sets of three samples, each packed in cylindrical containers with known activities of ²¹⁰Pb and ¹³⁷Cs, according to the preparation procedure using standard solutions. The sample filled with resin was meant to be considered as a standard source by the participants and could be used for the efficiency calibration of the detectors. The goal of the exercise was to determine the activity of the two radionuclides included in the other two samples, with unknown matrices. This required the participants to establish the self-attenuation correction factors between the calibration and the measurement matrices. Inactive matrices can be used to experimentally determine the attenuation coefficients for calculating self-absorption.

The different approaches and results obtained by the participants are presented and general recommendations are proposed.

Corresponding author: marie-christine.lepy@cea.fr

Proposed session: Gamma-ray spectrometry

Presentation type: